

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Nobis Group

SDG No.: Q2747

Contract: NOBI03

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI	Antimony	1.83	2.0	92	70 - 130	P	08/05/2025	14:14	LB136712
	Arsenic	1.02	1.0	102	70 - 130	P	08/05/2025	14:14	LB136712
	Barium	8.77	10.0	88	70 - 130	P	08/05/2025	14:14	LB136712
	Beryllium	1.07	1.0	107	70 - 130	P	08/05/2025	14:14	LB136712
	Cadmium	0.86	1.0	86	70 - 130	P	08/05/2025	14:14	LB136712
	Chromium	1.84	2.0	92	70 - 130	P	08/05/2025	14:14	LB136712
	Copper	1.70	2.0	85	70 - 130	P	08/05/2025	14:14	LB136712
	Lead	0.88	1.0	88	70 - 130	P	08/05/2025	14:14	LB136712
	Nickel	1.05	1.0	105	70 - 130	P	08/05/2025	14:14	LB136712
	Selenium	4.70	5.0	94	70 - 130	P	08/05/2025	14:14	LB136712
	Silver	0.92	1.0	92	70 - 130	P	08/05/2025	14:14	LB136712
	Thallium	0.88	1.0	88	70 - 130	P	08/05/2025	14:14	LB136712
	Vanadium	4.63	5.0	93	70 - 130	P	08/05/2025	14:14	LB136712
	Zinc	4.35	5.0	87	50 - 150	P	08/05/2025	14:14	LB136712
CRA	Mercury	0.16	0.2	81	70 - 130	CV	08/07/2025	09:58	LB136732
CRI	Antimony	2.08	2.0	104	70 - 130	P	08/12/2025	11:23	LB136790
	Arsenic	0.97	1.0	97	70 - 130	P	08/12/2025	11:23	LB136790
	Barium	9.63	10.0	96	70 - 130	P	08/12/2025	11:23	LB136790
	Beryllium	1.16	1.0	116	70 - 130	P	08/12/2025	11:23	LB136790
	Cadmium	0.79	1.0	79	70 - 130	P	08/12/2025	11:23	LB136790
	Chromium	2.02	2.0	101	70 - 130	P	08/12/2025	11:23	LB136790
	Copper	1.94	2.0	97	70 - 130	P	08/12/2025	11:23	LB136790
	Lead	0.85	1.0	85	70 - 130	P	08/12/2025	11:23	LB136790
	Nickel	0.82	1.0	82	70 - 130	P	08/12/2025	11:23	LB136790
	Selenium	5.28	5.0	106	70 - 130	P	08/12/2025	11:23	LB136790
	Silver	1.10	1.0	110	70 - 130	P	08/12/2025	11:23	LB136790
	Thallium	0.98	1.0	98	70 - 130	P	08/12/2025	11:23	LB136790
	Vanadium	5.09	5.0	102	70 - 130	P	08/12/2025	11:23	LB136790
	Zinc	4.48	5.0	90	50 - 150	P	08/12/2025	11:23	LB136790