



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2747

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB56	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	08/07/2025	09:46	LB136732

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2747

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB95	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	08/07/2025	09:51	LB136732
CCB96	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	08/07/2025	10:26	LB136732
CCB97	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	08/07/2025	10:44	LB136732

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2747

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Antimony	0.11	+/-1	U	0.25	2.00	P	08/05/2025	13:12	LB136712
	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	08/05/2025	13:12	LB136712
	Barium	0.21	+/-5	U	1.25	10.0	P	08/05/2025	13:12	LB136712
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	08/05/2025	13:12	LB136712
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	08/05/2025	13:12	LB136712
	Chromium	0.21	+/-1	U	0.75	2.00	P	08/05/2025	13:12	LB136712
	Copper	0.30	+/-1	U	1.50	2.00	P	08/05/2025	13:12	LB136712
	Lead	0.21	+/-0.5	U	0.75	1.00	P	08/05/2025	13:12	LB136712
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	08/05/2025	13:12	LB136712
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	08/05/2025	13:12	LB136712
	Silver	0.060	+/-0.5	U	0.50	1.00	P	08/05/2025	13:12	LB136712
	Thallium	0.060	+/-0.5	U	0.50	1.00	P	08/05/2025	13:12	LB136712
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	08/05/2025	13:12	LB136712
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	08/05/2025	13:12	LB136712

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2747

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Antimony	0.14	+/-1	J	0.25	2.00	P	08/05/2025	14:10	LB136712
	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	08/05/2025	14:10	LB136712
	Barium	0.21	+/-5	U	1.25	10.0	P	08/05/2025	14:10	LB136712
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	08/05/2025	14:10	LB136712
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	08/05/2025	14:10	LB136712
	Chromium	0.21	+/-1	U	0.75	2.00	P	08/05/2025	14:10	LB136712
	Copper	0.30	+/-1	U	1.50	2.00	P	08/05/2025	14:10	LB136712
	Lead	0.21	+/-0.5	U	0.75	1.00	P	08/05/2025	14:10	LB136712
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	08/05/2025	14:10	LB136712
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	08/05/2025	14:10	LB136712
	Silver	0.060	+/-0.5	U	0.50	1.00	P	08/05/2025	14:10	LB136712
	Thallium	0.10	+/-0.5	J	0.50	1.00	P	08/05/2025	14:10	LB136712
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	08/05/2025	14:10	LB136712
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	08/05/2025	14:10	LB136712
CCB02	Antimony	0.11	+/-1	U	0.25	2.00	P	08/05/2025	14:53	LB136712
	Arsenic	0.16	+/-0.5	J	0.25	1.00	P	08/05/2025	14:53	LB136712
	Barium	0.21	+/-5	U	1.25	10.0	P	08/05/2025	14:53	LB136712
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	08/05/2025	14:53	LB136712
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	08/05/2025	14:53	LB136712
	Chromium	0.21	+/-1	U	0.75	2.00	P	08/05/2025	14:53	LB136712
	Copper	0.30	+/-1	U	1.50	2.00	P	08/05/2025	14:53	LB136712
	Lead	0.21	+/-0.5	U	0.75	1.00	P	08/05/2025	14:53	LB136712
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	08/05/2025	14:53	LB136712
	Selenium	3.05	+/-2.5	J*	4.50	5.00	P	08/05/2025	14:53	LB136712
	Silver	0.060	+/-0.5	U	0.50	1.00	P	08/05/2025	14:53	LB136712
	Thallium	0.060	+/-0.5	U	0.50	1.00	P	08/05/2025	14:53	LB136712
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	08/05/2025	14:53	LB136712
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	08/05/2025	14:53	LB136712
CCB03	Antimony	0.14	+/-1	J	0.25	2.00	P	08/05/2025	15:30	LB136712
	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	08/05/2025	15:30	LB136712
	Barium	0.21	+/-5	U	1.25	10.0	P	08/05/2025	15:30	LB136712
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	08/05/2025	15:30	LB136712
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	08/05/2025	15:30	LB136712
	Chromium	0.21	+/-1	U	0.75	2.00	P	08/05/2025	15:30	LB136712
	Copper	0.30	+/-1	U	1.50	2.00	P	08/05/2025	15:30	LB136712
	Lead	0.21	+/-0.5	U	0.75	1.00	P	08/05/2025	15:30	LB136712
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	08/05/2025	15:30	LB136712
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	08/05/2025	15:30	LB136712
	Silver	0.060	+/-0.5	U	0.50	1.00	P	08/05/2025	15:30	LB136712

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2747

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB03	Thallium	0.060	+/-0.5	J	0.50	1.00	P	08/05/2025	15:30	LB136712
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	08/05/2025	15:30	LB136712
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	08/05/2025	15:30	LB136712
CCB04	Antimony	0.15	+/-1	J	0.25	2.00	P	08/05/2025	16:08	LB136712
	Arsenic	0.37	+/-0.5	J	0.25	1.00	P	08/05/2025	16:08	LB136712
	Barium	0.21	+/-5	U	1.25	10.0	P	08/05/2025	16:08	LB136712
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	08/05/2025	16:08	LB136712
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	08/05/2025	16:08	LB136712
	Chromium	0.21	+/-1	U	0.75	2.00	P	08/05/2025	16:08	LB136712
	Copper	0.30	+/-1	U	1.50	2.00	P	08/05/2025	16:08	LB136712
	Lead	0.21	+/-0.5	U	0.75	1.00	P	08/05/2025	16:08	LB136712
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	08/05/2025	16:08	LB136712
	Selenium	3.92	+/-2.5	J*	4.50	5.00	P	08/05/2025	16:08	LB136712
	Silver	0.060	+/-0.5	U	0.50	1.00	P	08/05/2025	16:08	LB136712
	Thallium	0.060	+/-0.5	U	0.50	1.00	P	08/05/2025	16:08	LB136712
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	08/05/2025	16:08	LB136712
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	08/05/2025	16:08	LB136712
	Antimony	0.11	+/-1	U	0.25	2.00	P	08/05/2025	17:02	LB136712
CCB05	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	08/05/2025	17:02	LB136712
	Barium	0.21	+/-5	U	1.25	10.0	P	08/05/2025	17:02	LB136712
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	08/05/2025	17:02	LB136712
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	08/05/2025	17:02	LB136712
	Chromium	0.21	+/-1	U	0.75	2.00	P	08/05/2025	17:02	LB136712
	Copper	0.30	+/-1	U	1.50	2.00	P	08/05/2025	17:02	LB136712
	Lead	0.21	+/-0.5	U	0.75	1.00	P	08/05/2025	17:02	LB136712
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	08/05/2025	17:02	LB136712
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	08/05/2025	17:02	LB136712
	Silver	0.060	+/-0.5	U	0.50	1.00	P	08/05/2025	17:02	LB136712
	Thallium	0.060	+/-0.5	U	0.50	1.00	P	08/05/2025	17:02	LB136712
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	08/05/2025	17:02	LB136712
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	08/05/2025	17:02	LB136712
	Antimony	0.11	+/-1	U	0.25	2.00	P	08/05/2025	17:47	LB136712
	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	08/05/2025	17:47	LB136712
CCB06	Barium	0.21	+/-5	U	1.25	10.0	P	08/05/2025	17:47	LB136712
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	08/05/2025	17:47	LB136712
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	08/05/2025	17:47	LB136712
	Chromium	0.21	+/-1	U	0.75	2.00	P	08/05/2025	17:47	LB136712
	Copper	0.30	+/-1	U	1.50	2.00	P	08/05/2025	17:47	LB136712
	Lead	0.21	+/-0.5	U	0.75	1.00	P	08/05/2025	17:47	LB136712

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2747

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Nickel	0.27	+/-0.5	U	0.75	1.00	P	08/05/2025	17:47	LB136712
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	08/05/2025	17:47	LB136712
	Silver	0.060	+/-0.5	U	0.50	1.00	P	08/05/2025	17:47	LB136712
	Thallium	0.060	+/-0.5	U	0.50	1.00	P	08/05/2025	17:47	LB136712
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	08/05/2025	17:47	LB136712
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	08/05/2025	17:47	LB136712
CCB07	Antimony	0.11	+/-1	U	0.25	2.00	P	08/05/2025	18:00	LB136712
	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	08/05/2025	18:00	LB136712
	Barium	0.21	+/-5	U	1.25	10.0	P	08/05/2025	18:00	LB136712
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	08/05/2025	18:00	LB136712
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	08/05/2025	18:00	LB136712
	Chromium	0.21	+/-1	U	0.75	2.00	P	08/05/2025	18:00	LB136712
	Copper	0.30	+/-1	U	1.50	2.00	P	08/05/2025	18:00	LB136712
	Lead	0.21	+/-0.5	U	0.75	1.00	P	08/05/2025	18:00	LB136712
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	08/05/2025	18:00	LB136712
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	08/05/2025	18:00	LB136712
	Silver	0.060	+/-0.5	U	0.50	1.00	P	08/05/2025	18:00	LB136712
	Thallium	0.060	+/-0.5	U	0.50	1.00	P	08/05/2025	18:00	LB136712
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	08/05/2025	18:00	LB136712
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	08/05/2025	18:00	LB136712

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2747

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Antimony	0.11	+/-1	U	0.25	2.00	P	08/12/2025	11:04	LB136790
	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	08/12/2025	11:04	LB136790
	Barium	0.21	+/-5	U	1.25	10.0	P	08/12/2025	11:04	LB136790
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	08/12/2025	11:04	LB136790
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	08/12/2025	11:04	LB136790
	Chromium	0.21	+/-1	U	0.75	2.00	P	08/12/2025	11:04	LB136790
	Copper	0.30	+/-1	U	1.50	2.00	P	08/12/2025	11:04	LB136790
	Lead	0.21	+/-0.5	U	0.75	1.00	P	08/12/2025	11:04	LB136790
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	08/12/2025	11:04	LB136790
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	08/12/2025	11:04	LB136790
	Silver	0.060	+/-0.5	U	0.50	1.00	P	08/12/2025	11:04	LB136790
	Thallium	0.060	+/-0.5	U	0.50	1.00	P	08/12/2025	11:04	LB136790
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	08/12/2025	11:04	LB136790
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	08/12/2025	11:04	LB136790

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2747

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Antimony	0.11	+/-1	U	0.25	2.00	P	08/12/2025	11:20	LB136790
	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	08/12/2025	11:20	LB136790
	Barium	0.21	+/-5	U	1.25	10.0	P	08/12/2025	11:20	LB136790
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	08/12/2025	11:20	LB136790
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	08/12/2025	11:20	LB136790
	Chromium	0.21	+/-1	U	0.75	2.00	P	08/12/2025	11:20	LB136790
	Copper	0.30	+/-1	U	1.50	2.00	P	08/12/2025	11:20	LB136790
	Lead	0.21	+/-0.5	U	0.75	1.00	P	08/12/2025	11:20	LB136790
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	08/12/2025	11:20	LB136790
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	08/12/2025	11:20	LB136790
	Silver	0.060	+/-0.5	U	0.50	1.00	P	08/12/2025	11:20	LB136790
	Thallium	0.060	+/-0.5	U	0.50	1.00	P	08/12/2025	11:20	LB136790
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	08/12/2025	11:20	LB136790
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	08/12/2025	11:20	LB136790
CCB02	Antimony	0.11	+/-1	U	0.25	2.00	P	08/12/2025	12:05	LB136790
	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	08/12/2025	12:05	LB136790
	Barium	0.21	+/-5	U	1.25	10.0	P	08/12/2025	12:05	LB136790
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	08/12/2025	12:05	LB136790
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	08/12/2025	12:05	LB136790
	Chromium	0.21	+/-1	U	0.75	2.00	P	08/12/2025	12:05	LB136790
	Copper	0.30	+/-1	U	1.50	2.00	P	08/12/2025	12:05	LB136790
	Lead	0.21	+/-0.5	U	0.75	1.00	P	08/12/2025	12:05	LB136790
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	08/12/2025	12:05	LB136790
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	08/12/2025	12:05	LB136790
	Silver	0.060	+/-0.5	U	0.50	1.00	P	08/12/2025	12:05	LB136790
	Thallium	0.060	+/-0.5	U	0.50	1.00	P	08/12/2025	12:05	LB136790
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	08/12/2025	12:05	LB136790
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	08/12/2025	12:05	LB136790
CCB03	Antimony	0.11	+/-1	U	0.25	2.00	P	08/12/2025	12:23	LB136790
	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	08/12/2025	12:23	LB136790
	Barium	0.21	+/-5	U	1.25	10.0	P	08/12/2025	12:23	LB136790
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	08/12/2025	12:23	LB136790
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	08/12/2025	12:23	LB136790
	Chromium	0.21	+/-1	U	0.75	2.00	P	08/12/2025	12:23	LB136790
	Copper	0.30	+/-1	U	1.50	2.00	P	08/12/2025	12:23	LB136790
	Lead	0.21	+/-0.5	U	0.75	1.00	P	08/12/2025	12:23	LB136790
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	08/12/2025	12:23	LB136790
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	08/12/2025	12:23	LB136790
	Silver	0.060	+/-0.5	U	0.50	1.00	P	08/12/2025	12:23	LB136790

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2747

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB03	Thallium	0.060	+/-0.5	U	0.50	1.00	P	08/12/2025	12:23	LB136790
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	08/12/2025	12:23	LB136790
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	08/12/2025	12:23	LB136790
CCB04	Antimony	0.11	+/-1	U	0.25	2.00	P	08/12/2025	12:47	LB136790
	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	08/12/2025	12:47	LB136790
	Barium	0.21	+/-5	U	1.25	10.0	P	08/12/2025	12:47	LB136790
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	08/12/2025	12:47	LB136790
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	08/12/2025	12:47	LB136790
	Chromium	0.21	+/-1	U	0.75	2.00	P	08/12/2025	12:47	LB136790
	Copper	0.30	+/-1	U	1.50	2.00	P	08/12/2025	12:47	LB136790
	Lead	0.21	+/-0.5	U	0.75	1.00	P	08/12/2025	12:47	LB136790
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	08/12/2025	12:47	LB136790
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	08/12/2025	12:47	LB136790
	Silver	0.060	+/-0.5	U	0.50	1.00	P	08/12/2025	12:47	LB136790
	Thallium	0.060	+/-0.5	U	0.50	1.00	P	08/12/2025	12:47	LB136790
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	08/12/2025	12:47	LB136790
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	08/12/2025	12:47	LB136790