

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Nobis Group **Level:** LOW **SDG No.:** Q2747
Contract: NOBI03 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q2747-10 **Client ID:** OU4-TS-GRILLO-TSCP02-073125DUP
Percent Solids for Sample: NA **Duplicate ID** Q2747-10DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Antimony	ug/L	20	10.0	UD	10.0	UD			P
Arsenic	ug/L	20	19.7	D	19.7	D	0		P
Barium	ug/L	20	113	D	110	D	3		P
Beryllium	ug/L	20	7.60	D	6.90	D	10		P
Cadmium	ug/L	20	7.65	D	7.20	D	6		P
Chromium	ug/L	20	11.5	D	11.3	D	2		P
Copper	ug/L	20	68.6	D	67.5	D	2		P
Lead	ug/L	20	48.8	D	47.8	D	2		P
Nickel	ug/L	20	45.5	D	46.0	D	1		P
Mercury	ug/L	20	0.20	U	0.20	U			CV
Selenium	ug/L	20	25.0	UD	25.0	UD			P
Silver	ug/L	20	5.00	UD	5.00	UD			P
Thallium	ug/L	20	5.00	UD	5.00	UD			P
Vanadium	ug/L	20	82.9	D	82.3	D	1		P
Zinc	ug/L	20	1680	D	1640	D	2		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Nobis Group **Level:** LOW **SDG No.:** Q2747
Contract: NOBI03 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q2747-10MS **Client ID:** OU4-TS-GRILLO-TSCP02-073125MSD
Percent Solids for Sample: NA **Duplicate ID** Q2747-10MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Antimony	ug/L	20	394	D	402	D	2		P
Arsenic	ug/L	20	160	D	270	D	51	*	P
Barium	ug/L	20	1820	D	1860	D	2		P
Beryllium	ug/L	20	463	D	459	D	1		P
Cadmium	ug/L	20	472	D	482	D	2		P
Chromium	ug/L	20	404	D	403	D	0		P
Copper	ug/L	20	4470	D	4500	D	1		P
Lead	ug/L	20	2370	D	2380	D	0		P
Nickel	ug/L	20	510	D	512	D	0		P
Mercury	ug/L	20	3.93		3.96		1		CV
Selenium	ug/L	20	425	D	429	D	1		P
Silver	ug/L	20	65.8	D	67.2	D	2		P
Thallium	ug/L	20	364	D	361	D	1		P
Vanadium	ug/L	20	485	D	485	D	0		P
Zinc	ug/L	20	5350	D	5370	D	0		P