

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Nobis Group

SDG No.: Q2747

Contract: NOBI03

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169128BS							
Antimony	ug/L	500	539		108	85 - 117	P
Arsenic	ug/L	500	518		104	84 - 116	P
Barium	ug/L	2500	2630		105	86 - 114	P
Beryllium	ug/L	500	535		107	83 - 121	P
Cadmium	ug/L	500	533		107	87 - 115	P
Chromium	ug/L	500	534		107	85 - 116	P
Copper	ug/L	5000	5350		107	85 - 118	P
Lead	ug/L	2500	2640		106	88 - 115	P
Nickel	ug/L	500	533		107	85 - 117	P
Selenium	ug/L	500	518		104	80 - 120	P
Silver	ug/L	500	545		109	85 - 116	P
Thallium	ug/L	500	530		106	82 - 116	P
Vanadium	ug/L	500	524		105	86 - 115	P
Zinc	ug/L	5000	5350		107	83 - 119	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Nobis Group

SDG No.: Q2747

Contract: NOBI03

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169158BS Mercury	ug/L	4.0	4.24		106	82 - 119	CV