

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: ACE

SDG No.: Q2747

Instrument ID: P7

Start Date : 08/05/2025

Run Number: LB136712

End Date : 08/05/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
S0	S0	1128	100		100							
S2	S2	1134	100		100							
S3	S3	1138	101		100							
S4	S4	1141	103		101							
S5	S5	1144	103		103							
S6	S6	1147	104		104							
S7	S7	1149	106		103							
S8	S8	1152	103		94							
ICV01	ICV01	1252	101		100							
LLICV01	LLICV01	1308	102		97							
ICB01	ICB01	1312	100		96							
ICSA01	ICSA01	1339	105		100							
ICSAB01	ICSAB01	1404	109		100							
CCV01	CCV01	1407	111		99							
CCB01	CCB01	1410	106		101							
CRI	CRI	1414	106		100							
ZZZZZZ	ZZZZZZ	1418										
ZZZZZZ	ZZZZZZ	1431										
ZZZZZZ	ZZZZZZ	1434										
ZZZZZZ	ZZZZZZ	1437										
ZZZZZZ	ZZZZZZ	1440										
ZZZZZZ	ZZZZZZ	1443										
CCV02	CCV02	1446	119		106							
CCB02	CCB02	1453	111		105							
ZZZZZZ	ZZZZZZ	1457										
ZZZZZZ	ZZZZZZ	1500										
ZZZZZZ	ZZZZZZ	1504										
ZZZZZZ	ZZZZZZ	1507										
ZZZZZZ	ZZZZZZ	1510										
ZZZZZZ	ZZZZZZ	1513										
ZZZZZZ	ZZZZZZ	1515										
ZZZZZZ	ZZZZZZ	1518										
ZZZZZZ	ZZZZZZ	1524										
CCV03	CCV03	1528	119		105							

Internal Standard %RI Limit: 30 -120

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: ACE

SDG No.: Q2747

Instrument ID: P7

Start Date : 08/05/2025

Run Number: LB136712

End Date : 08/05/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
CCB03	CCB03	1530	113		106							
ZZZZZZ	ZZZZZZ	1534										
ZZZZZZ	ZZZZZZ	1537										
ZZZZZZ	ZZZZZZ	1540										
ZZZZZZ	ZZZZZZ	1543										
ZZZZZZ	ZZZZZZ	1546										
ZZZZZZ	ZZZZZZ	1556										
ZZZZZZ	ZZZZZZ	1559										
ZZZZZZ	ZZZZZZ	1602										
CCV04	CCV04	1605	118		104							
CCB04	CCB04	1608	115		106							
ZZZZZZ	ZZZZZZ	1613										
ZZZZZZ	ZZZZZZ	1616										
ZZZZZZ	ZZZZZZ	1619										
ZZZZZZ	ZZZZZZ	1622										
ZZZZZZ	ZZZZZZ	1625										
ZZZZZZ	ZZZZZZ	1628										
ZZZZZZ	ZZZZZZ	1630										
ZZZZZZ	ZZZZZZ	1634										
CCV05	CCV05	1643	115		102							
CCB05	CCB05	1702	110		102							
PB169128TB	PB169128TB	1705	110		103							
Q2747-02	OU4-TS-GRILI	1709	115		105							
Q2747-04	OU4-TS-GRILI	1712	115		104							
Q2747-06	OU4-TS-GRILI	1715	116		104							
Q2747-08	OU4-TS-GRILI	1718	114		103							
Q2747-10	OU4-TS-GRILI	1721	116		105							
Q2747-10DUP	OU4-TS-GRILI	1725	117		104							
Q2747-10L	OU4-TS-GRILI	1728	116		103							
Q2747-10MS	OU4-TS-GRILI	1731	115		103							
Q2747-10MSD	OU4-TS-GRILI	1734	116		105							
CCV06	CCV06	1737	118		102							
CCB06	CCB06	1747	112		104							
Q2747-10A	OU4-TS-GRILI	1750	116		105							

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: ACE

SDG No.: Q2747

Instrument ID: P7

Start Date : 08/05/2025

Run Number: LB136712

End Date : 08/05/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
CCV07	CCV07	1753	116		101							
CCB07	CCB07	1800	111		103							

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: ACE

SDG No.: Q2747

Instrument ID: P7

Start Date : 08/05/2025

Run Number: LB136712

End Date : 08/05/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
S0	S0	1128	100									
S2	S2	1134	101									
S3	S3	1138	101									
S4	S4	1141	100									
S5	S5	1144	99									
S6	S6	1147	101									
S7	S7	1149	101									
S8	S8	1152	94									
ICV01	ICV01	1252	97									
LLICV01	LLICV01	1308	97									
ICB01	ICB01	1312	99									
ICSA01	ICSA01	1339	101									
ICSAB01	ICSAB01	1404	101									
CCV01	CCV01	1407	101									
CCB01	CCB01	1410	100									
CRI	CRI	1414	101									
ZZZZZZ	ZZZZZZ	1418										
ZZZZZZ	ZZZZZZ	1431										
ZZZZZZ	ZZZZZZ	1434										
ZZZZZZ	ZZZZZZ	1437										
ZZZZZZ	ZZZZZZ	1440										
ZZZZZZ	ZZZZZZ	1443										
CCV02	CCV02	1446	104									
CCB02	CCB02	1453	106									
ZZZZZZ	ZZZZZZ	1457										
ZZZZZZ	ZZZZZZ	1500										
ZZZZZZ	ZZZZZZ	1504										
ZZZZZZ	ZZZZZZ	1507										
ZZZZZZ	ZZZZZZ	1510										
ZZZZZZ	ZZZZZZ	1513										
ZZZZZZ	ZZZZZZ	1515										
ZZZZZZ	ZZZZZZ	1518										
ZZZZZZ	ZZZZZZ	1524										
CCV03	CCV03	1528	105									

Internal Standard %RI Limit: 30 -120

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: ACE

SDG No.: Q2747

Instrument ID: P7

Start Date : 08/05/2025

Run Number: LB136712

End Date : 08/05/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
CCB03	CCB03	1530	107									
ZZZZZZ	ZZZZZZ	1534										
ZZZZZZ	ZZZZZZ	1537										
ZZZZZZ	ZZZZZZ	1540										
ZZZZZZ	ZZZZZZ	1543										
ZZZZZZ	ZZZZZZ	1546										
ZZZZZZ	ZZZZZZ	1556										
ZZZZZZ	ZZZZZZ	1559										
ZZZZZZ	ZZZZZZ	1602										
CCV04	CCV04	1605	104									
CCB04	CCB04	1608	108									
ZZZZZZ	ZZZZZZ	1613										
ZZZZZZ	ZZZZZZ	1616										
ZZZZZZ	ZZZZZZ	1619										
ZZZZZZ	ZZZZZZ	1622										
ZZZZZZ	ZZZZZZ	1625										
ZZZZZZ	ZZZZZZ	1628										
ZZZZZZ	ZZZZZZ	1630										
ZZZZZZ	ZZZZZZ	1634										
CCV05	CCV05	1643	100									
CCB05	CCB05	1702	104									
PB169128TB	PB169128TB	1705	104									
Q2747-02	OU4-TS-GRILI	1709	108									
Q2747-04	OU4-TS-GRILI	1712	108									
Q2747-06	OU4-TS-GRILI	1715	107									
Q2747-08	OU4-TS-GRILI	1718	108									
Q2747-10	OU4-TS-GRILI	1721	109									
Q2747-10DUP	OU4-TS-GRILI	1725	108									
Q2747-10L	OU4-TS-GRILI	1728	107									
Q2747-10MS	OU4-TS-GRILI	1731	107									
Q2747-10MSD	OU4-TS-GRILI	1734	107									
CCV06	CCV06	1737	102									
CCB06	CCB06	1747	104									
Q2747-10A	OU4-TS-GRILI	1750	110									

Internal Standard %RI Limit: 30 -120

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: ACE

SDG No.: Q2747

Instrument ID: P7

Start Date : 08/05/2025

Run Number: LB136712

End Date : 08/05/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
CCV07	CCV07	1753	101									
CCB07	CCB07	1800	106									

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: ACE

SDG No.: Q2747

Instrument ID: P7

Start Date : 08/12/2025

Run Number: LB136790

End Date : 08/12/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
S0	S0	1008	100		100							
S2	S2	1011	97		98							
S3	S3	1018	100		99							
S4	S4	1021	98		100							
S5	S5	1024	100		100							
S6	S6	1027	101		100							
S7	S7	1029	102		99							
S8	S8	1032	98		90							
ICV01	ICV01	1051	100		99							
LLICV01	LLICV01	1054	99		99							
ICB01	ICB01	1104	99		98							
ICSA01	ICSA01	1111	103		97							
ICSAB01	ICSAB01	1114	107		100							
CCV01	CCV01	1117	103		95							
CCB01	CCB01	1120	102		101							
CRI	CRI	1123	102		100							
PB169128BL	PB169128BL	1128	102		99							
PB169128BS	PB169128BS	1131	101		97							
ZZZZZZ	ZZZZZZ	1134										
ZZZZZZ	ZZZZZZ	1137										
ZZZZZZ	ZZZZZZ	1140										
ZZZZZZ	ZZZZZZ	1143										
ZZZZZZ	ZZZZZZ	1146										
ZZZZZZ	ZZZZZZ	1149										
CCV02	CCV02	1201	104		96							
CCB02	CCB02	1205	103		98							

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: ACE

SDG No.: Q2747

Instrument ID: P7

Start Date : 08/12/2025

Run Number: LB136790

End Date : 08/12/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
S0	S0	1008	100									
S2	S2	1011	99									
S3	S3	1018	99									
S4	S4	1021	99									
S5	S5	1024	99									
S6	S6	1027	98									
S7	S7	1029	98									
S8	S8	1032	90									
ICV01	ICV01	1051	101									
LLICV01	LLICV01	1054	99									
ICB01	ICB01	1104	98									
ICSA01	ICSA01	1111	98									
ICSAB01	ICSAB01	1114	98									
CCV01	CCV01	1117	94									
CCB01	CCB01	1120	101									
CRI	CRI	1123	101									
PB169128BL	PB169128BL	1128	101									
PB169128BS	PB169128BS	1131	97									
ZZZZZZ	ZZZZZZ	1134										
ZZZZZZ	ZZZZZZ	1137										
ZZZZZZ	ZZZZZZ	1140										
ZZZZZZ	ZZZZZZ	1143										
ZZZZZZ	ZZZZZZ	1146										
ZZZZZZ	ZZZZZZ	1149										
CCV02	CCV02	1201	97									
CCB02	CCB02	1205	101									