

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2747
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/04/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031353.D **Time Analyzed:** 10:10

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.604	9.521	9.721	845.560	712.500	18.7
2,4,5-TP(Silvex)	9.310	9.227	9.427	777.450	712.500	9.1
2,4-D	8.431	8.346	8.546	788.230	705.000	11.8
2,4-DB	10.181	10.099	10.299	818.290	712.500	14.8
2,4-DCAA	7.302	7.215	7.415	790.880	750.000	5.5
Dalapon	2.681	2.591	2.791	666.870	682.500	-2.3
DICAMBA	7.489	7.403	7.603	718.060	705.000	1.9
DICHLORPROP	8.198	8.113	8.313	712.890	705.000	1.1
Dinoseb	11.386	11.307	11.507	761.110	705.000	8.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2747
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/04/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031353.D **Time Analyzed:** 10:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.339	10.252	10.452	669.110	712.500	-6.1
2,4,5-TP(Silvex)	9.911	9.824	10.024	671.720	712.500	-5.7
2,4-D	9.011	8.925	9.125	656.660	705.000	-6.9
2,4-DB	10.907	10.822	11.022	642.160	712.500	-9.9
2,4-DCAA	7.754	7.666	7.866	692.250	750.000	-7.7
Dalapon	2.698	2.603	2.803	629.750	682.500	-7.7
DICAMBA	7.953	7.865	8.065	658.370	705.000	-6.6
DICHLORPROP	8.672	8.585	8.785	644.590	705.000	-8.6
Dinoseb	11.287	11.202	11.402	636.760	705.000	-9.7