

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2747
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/04/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031358.D **Time Analyzed:** 18:11

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.604	9.521	9.721	753.780	712.500	5.8
2,4,5-TP(Silvex)	9.310	9.227	9.427	705.390	712.500	-1.0
2,4-D	8.431	8.346	8.546	715.860	705.000	1.5
2,4-DB	10.181	10.099	10.299	766.710	712.500	7.6
2,4-DCAA	7.302	7.215	7.415	728.930	750.000	-2.8
Dalapon	2.681	2.591	2.791	611.150	682.500	-10.5
DICAMBA	7.488	7.403	7.603	660.430	705.000	-6.3
DICHLORPROP	8.198	8.113	8.313	652.500	705.000	-7.4
Dinoseb	11.385	11.307	11.507	696.960	705.000	-1.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2747
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/04/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031358.D **Time Analyzed:** 18:11

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.339	10.252	10.452	620.420	712.500	-12.9
2,4,5-TP(Silvex)	9.911	9.824	10.024	624.120	712.500	-12.4
2,4-D	9.012	8.925	9.125	611.990	705.000	-13.2
2,4-DB	10.907	10.822	11.022	603.590	712.500	-15.3
2,4-DCAA	7.755	7.666	7.866	642.540	750.000	-14.3
Dalapon	2.700	2.603	2.803	585.170	682.500	-14.3
DICAMBA	7.954	7.865	8.065	608.820	705.000	-13.6
DICHLORPROP	8.673	8.585	8.785	599.420	705.000	-15.0
Dinoseb	11.287	11.202	11.402	605.920	705.000	-14.1