

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2747
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/04/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031368.D **Time Analyzed:** 22:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.602	9.521	9.721	772.840	712.500	8.5
2,4,5-TP(Silvex)	9.307	9.227	9.427	706.090	712.500	-0.9
2,4-D	8.429	8.346	8.546	721.970	705.000	2.4
2,4-DB	10.177	10.099	10.299	765.030	712.500	7.4
2,4-DCAA	7.300	7.215	7.415	723.560	750.000	-3.5
Dalapon	2.684	2.591	2.791	603.110	682.500	-11.6
DICAMBA	7.486	7.403	7.603	654.160	705.000	-7.2
DICHLORPROP	8.195	8.113	8.313	652.280	705.000	-7.5
Dinoseb	11.383	11.307	11.507	681.050	705.000	-3.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2747
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/04/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031368.D **Time Analyzed:** 22:13

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.337	10.252	10.452	632.240	712.500	-11.3
2,4,5-TP(Silvex)	9.909	9.824	10.024	630.470	712.500	-11.5
2,4-D	9.011	8.925	9.125	621.180	705.000	-11.9
2,4-DB	10.905	10.822	11.022	612.940	712.500	-14.0
2,4-DCAA	7.753	7.666	7.866	651.950	750.000	-13.1
Dalapon	2.701	2.603	2.803	593.390	682.500	-13.1
DICAMBA	7.953	7.865	8.065	618.890	705.000	-12.2
DICHLORPROP	8.672	8.585	8.785	610.630	705.000	-13.4
Dinoseb	11.285	11.202	11.402	599.070	705.000	-15.0