

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2747
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

Client Sample No.: CCAL04 **Date Analyzed:** 08/05/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031371.D **Time Analyzed:** 11:25

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.600	9.521	9.721	782.960	712.500	9.9
2,4,5-TP(Silvex)	9.306	9.227	9.427	725.140	712.500	1.8
2,4-D	8.427	8.346	8.546	732.140	705.000	3.8
2,4-DB	10.176	10.099	10.299	786.850	712.500	10.4
2,4-DCAA	7.300	7.215	7.415	742.360	750.000	-1.0
Dalapon	2.682	2.591	2.791	615.530	682.500	-9.8
DICAMBA	7.486	7.403	7.603	675.700	705.000	-4.2
DICHLORPROP	8.195	8.113	8.313	670.280	705.000	-4.9
Dinoseb	11.381	11.307	11.507	696.680	705.000	-1.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2747
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

Client Sample No.: CCAL04 **Date Analyzed:** 08/05/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031371.D **Time Analyzed:** 11:25

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.336	10.252	10.452	626.270	712.500	-12.1
2,4,5-TP(Silvex)	9.907	9.824	10.024	634.630	712.500	-10.9
2,4-D	9.009	8.925	9.125	618.420	705.000	-12.3
2,4-DB	10.903	10.822	11.022	609.190	712.500	-14.5
2,4-DCAA	7.752	7.666	7.866	650.050	750.000	-13.3
Dalapon	2.701	2.603	2.803	586.080	682.500	-14.1
DICAMBA	7.950	7.865	8.065	618.550	705.000	-12.3
DICHLORPROP	8.670	8.585	8.785	613.680	705.000	-13.0
Dinoseb	11.283	11.202	11.402	600.580	705.000	-14.8