

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2747
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

Client Sample No.: CCAL05 **Date Analyzed:** 08/05/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031376.D **Time Analyzed:** 15:48

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.600	9.521	9.721	801.640	712.500	12.5
2,4,5-TP(Silvex)	9.305	9.227	9.427	724.570	712.500	1.7
2,4-D	8.427	8.346	8.546	741.600	705.000	5.2
2,4-DB	10.175	10.099	10.299	835.150	712.500	17.2
2,4-DCAA	7.299	7.215	7.415	730.290	750.000	-2.6
Dalapon	2.680	2.591	2.791	606.150	682.500	-11.2
DICAMBA	7.485	7.403	7.603	666.730	705.000	-5.4
DICHLORPROP	8.194	8.113	8.313	666.530	705.000	-5.5
Dinoseb	11.381	11.307	11.507	671.020	705.000	-4.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2747
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

Client Sample No.: CCAL05 **Date Analyzed:** 08/05/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031376.D **Time Analyzed:** 15:48

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.335	10.252	10.452	646.960	712.500	-9.2
2,4,5-TP(Silvex)	9.907	9.824	10.024	646.110	712.500	-9.3
2,4-D	9.008	8.925	9.125	619.270	705.000	-12.2
2,4-DB	10.903	10.822	11.022	630.300	712.500	-11.5
2,4-DCAA	7.752	7.666	7.866	663.940	750.000	-11.5
Dalapon	2.699	2.603	2.803	599.420	682.500	-12.2
DICAMBA	7.951	7.865	8.065	631.600	705.000	-10.4
DICHLORPROP	8.670	8.585	8.785	598.680	705.000	-15.1
Dinoseb	11.282	11.202	11.402	583.140	705.000	-17.3