

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: NOBI03

Lab Code: ACE

SDG NO.: Q2747

Continuing Calib Date: 08/04/2025

Initial Calibration Date(s): 07/23/2025

07/23/2025

Continuing Calib Time: 18:54

Initial Calibration Time(s): 11:32

19:47

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.75       | 4.76      | 4.66      | 4.86 | 0.01       |
| Aroclor-1016-2       | (2) | 4.77       | 4.77      | 4.67      | 4.87 | 0.00       |
| Aroclor-1016-3       | (3) | 4.83       | 4.83      | 4.73      | 4.93 | 0.00       |
| Aroclor-1016-4       | (4) | 4.95       | 4.95      | 4.85      | 5.05 | 0.00       |
| Aroclor-1016-5       | (5) | 5.20       | 5.21      | 5.11      | 5.31 | 0.01       |
| Aroclor-1260-1       | (1) | 6.24       | 6.25      | 6.15      | 6.35 | 0.01       |
| Aroclor-1260-2       | (2) | 6.43       | 6.43      | 6.33      | 6.53 | 0.00       |
| Aroclor-1260-3       | (3) | 6.80       | 6.80      | 6.70      | 6.90 | 0.00       |
| Aroclor-1260-4       | (4) | 7.05       | 7.06      | 6.96      | 7.16 | 0.01       |
| Aroclor-1260-5       | (5) | 7.30       | 7.30      | 7.20      | 7.40 | 0.00       |
| Tetrachloro-m-xylene |     | 3.67       | 3.67      | 3.57      | 3.77 | 0.00       |
| Decachlorobiphenyl   |     | 8.69       | 8.69      | 8.59      | 8.79 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: NOBI03

Lab Code: ACE

SDG NO.: Q2747

Continuing Calib Date: 08/04/2025

Initial Calibration Date(s): 07/23/2025

07/23/2025

Continuing Calib Time: 18:54

Initial Calibration Time(s): 11:32

19:47

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.73       | 4.74      | 4.64      | 4.84 | 0.01       |
| Aroclor-1016-2       | (2) | 4.75       | 4.76      | 4.66      | 4.86 | 0.01       |
| Aroclor-1016-3       | (3) | 4.93       | 4.93      | 4.83      | 5.03 | 0.00       |
| Aroclor-1016-4       | (4) | 4.97       | 4.97      | 4.87      | 5.07 | 0.00       |
| Aroclor-1016-5       | (5) | 5.18       | 5.19      | 5.09      | 5.29 | 0.01       |
| Aroclor-1260-1       | (1) | 6.21       | 6.21      | 6.11      | 6.31 | 0.00       |
| Aroclor-1260-2       | (2) | 6.40       | 6.40      | 6.30      | 6.50 | 0.00       |
| Aroclor-1260-3       | (3) | 6.55       | 6.55      | 6.45      | 6.65 | 0.00       |
| Aroclor-1260-4       | (4) | 7.02       | 7.02      | 6.92      | 7.12 | 0.00       |
| Aroclor-1260-5       | (5) | 7.26       | 7.26      | 7.16      | 7.36 | 0.00       |
| Tetrachloro-m-xylene |     | 3.66       | 3.66      | 3.56      | 3.76 | 0.00       |
| Decachlorobiphenyl   |     | 8.63       | 8.64      | 8.54      | 8.74 | 0.01       |