

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: NOBI03

Lab Code: ACE

SDG NO.: Q2747

Continuing Calib Date: 08/05/2025

Initial Calibration Date(s): 07/23/2025

07/23/2025

Continuing Calib Time: 16:34

Initial Calibration Time(s): 11:32

19:47

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.75	4.75	4.65	4.85	0.00
Aroclor-1248-2 (2)	4.99	4.99	4.89	5.09	0.00
Aroclor-1248-3 (3)	5.20	5.21	5.11	5.31	0.01
Aroclor-1248-4 (4)	5.56	5.56	5.46	5.66	0.00
Aroclor-1248-5 (5)	5.60	5.60	5.50	5.70	0.00
Tetrachloro-m-xylene	3.67	3.67	3.57	3.77	0.00
Decachlorobiphenyl	8.69	8.69	8.59	8.79	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: NOBI03

Lab Code: ACE

SDG NO.: Q2747

Continuing Calib Date: 08/05/2025

Initial Calibration Date(s): 07/23/2025

07/23/2025

Continuing Calib Time: 16:34

Initial Calibration Time(s): 11:32

19:47

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.73	4.74	4.64	4.84	0.01
Aroclor-1248-2 (2)	4.97	4.97	4.87	5.07	0.00
Aroclor-1248-3 (3)	5.01	5.02	4.92	5.12	0.01
Aroclor-1248-4 (4)	5.18	5.19	5.09	5.29	0.01
Aroclor-1248-5 (5)	5.57	5.58	5.48	5.68	0.01
Tetrachloro-m-xylene	3.66	3.66	3.56	3.76	0.00
Decachlorobiphenyl	8.63	8.64	8.54	8.74	0.01