

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2747
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/04/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112718.D **Time Analyzed:** 12:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.750	4.655	4.855	611.200	500.000	22.2
Aroclor-1016-2	4.769	4.674	4.874	539.760	500.000	8.0
Aroclor-1016-3	4.827	4.731	4.931	541.820	500.000	8.4
Aroclor-1016-4	4.946	4.851	5.051	552.060	500.000	10.4
Aroclor-1016-5	5.202	5.108	5.308	578.860	500.000	15.8
Aroclor-1260-1	6.240	6.145	6.345	553.900	500.000	10.8
Aroclor-1260-2	6.429	6.333	6.533	547.200	500.000	9.4
Aroclor-1260-3	6.795	6.700	6.900	577.380	500.000	15.5
Aroclor-1260-4	7.055	6.959	7.159	553.500	500.000	10.7
Aroclor-1260-5	7.297	7.203	7.403	554.550	500.000	10.9
Decachlorobiphenyl	8.686	8.593	8.793	52.570	50.000	5.1
Tetrachloro-m-xylene	3.666	3.569	3.769	56.010	50.000	12.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2747
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/04/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112718.D **Time Analyzed:** 12:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.734	4.638	4.838	513.580	500.000	2.7
Aroclor-1016-2	4.752	4.656	4.856	504.760	500.000	1.0
Aroclor-1016-3	4.927	4.831	5.031	505.790	500.000	1.2
Aroclor-1016-4	4.969	4.873	5.073	507.940	500.000	1.6
Aroclor-1016-5	5.181	5.085	5.285	515.790	500.000	3.2
Aroclor-1260-1	6.209	6.113	6.313	524.320	500.000	4.9
Aroclor-1260-2	6.397	6.302	6.502	529.280	500.000	5.9
Aroclor-1260-3	6.548	6.453	6.653	559.700	500.000	11.9
Aroclor-1260-4	7.017	6.923	7.123	531.590	500.000	6.3
Aroclor-1260-5	7.260	7.164	7.364	557.900	500.000	11.6
Decachlorobiphenyl	8.634	8.540	8.740	52.550	50.000	5.1
Tetrachloro-m-xylene	3.660	3.562	3.762	52.730	50.000	5.5