

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2747
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/04/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112728.D **Time Analyzed:** 16:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.753	4.655	4.855	542.310	500.000	8.5
Aroclor-1016-2	4.771	4.674	4.874	544.310	500.000	8.9
Aroclor-1016-3	4.828	4.731	4.931	531.700	500.000	6.3
Aroclor-1016-4	4.948	4.851	5.051	537.730	500.000	7.5
Aroclor-1016-5	5.204	5.108	5.308	585.920	500.000	17.2
Aroclor-1260-1	6.241	6.145	6.345	531.610	500.000	6.3
Aroclor-1260-2	6.431	6.333	6.533	535.830	500.000	7.2
Aroclor-1260-3	6.797	6.700	6.900	556.820	500.000	11.4
Aroclor-1260-4	7.056	6.959	7.159	524.850	500.000	5.0
Aroclor-1260-5	7.300	7.203	7.403	525.610	500.000	5.1
Decachlorobiphenyl	8.688	8.593	8.793	50.400	50.000	0.8
Tetrachloro-m-xylene	3.667	3.569	3.769	54.350	50.000	8.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2747
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/04/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112728.D **Time Analyzed:** 16:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.734	4.638	4.838	518.090	500.000	3.6
Aroclor-1016-2	4.752	4.656	4.856	507.610	500.000	1.5
Aroclor-1016-3	4.926	4.831	5.031	505.890	500.000	1.2
Aroclor-1016-4	4.969	4.873	5.073	503.770	500.000	0.8
Aroclor-1016-5	5.181	5.085	5.285	524.580	500.000	4.9
Aroclor-1260-1	6.209	6.113	6.313	530.550	500.000	6.1
Aroclor-1260-2	6.397	6.302	6.502	529.660	500.000	5.9
Aroclor-1260-3	6.548	6.453	6.653	550.420	500.000	10.1
Aroclor-1260-4	7.018	6.923	7.123	537.980	500.000	7.6
Aroclor-1260-5	7.260	7.164	7.364	555.820	500.000	11.2
Decachlorobiphenyl	8.634	8.540	8.740	52.460	50.000	4.9
Tetrachloro-m-xylene	3.660	3.562	3.762	52.840	50.000	5.7