

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2747
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/05/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112740.D **Time Analyzed:** 08:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.751	4.655	4.855	588.790	500.000	17.8
Aroclor-1016-2	4.770	4.674	4.874	588.650	500.000	17.7
Aroclor-1016-3	4.828	4.731	4.931	570.570	500.000	14.1
Aroclor-1016-4	4.946	4.851	5.051	579.010	500.000	15.8
Aroclor-1016-5	5.203	5.108	5.308	591.820	500.000	18.4
Aroclor-1260-1	6.240	6.145	6.345	583.750	500.000	16.8
Aroclor-1260-2	6.429	6.333	6.533	575.540	500.000	15.1
Aroclor-1260-3	6.795	6.700	6.900	589.860	500.000	18.0
Aroclor-1260-4	7.054	6.959	7.159	575.960	500.000	15.2
Aroclor-1260-5	7.297	7.203	7.403	570.460	500.000	14.1
Decachlorobiphenyl	8.686	8.593	8.793	55.110	50.000	10.2
Tetrachloro-m-xylene	3.667	3.569	3.769	59.130	50.000	18.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2747
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/05/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112740.D **Time Analyzed:** 08:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.732	4.638	4.838	534.220	500.000	6.8
Aroclor-1016-2	4.751	4.656	4.856	533.520	500.000	6.7
Aroclor-1016-3	4.926	4.831	5.031	526.410	500.000	5.3
Aroclor-1016-4	4.968	4.873	5.073	523.300	500.000	4.7
Aroclor-1016-5	5.180	5.085	5.285	550.320	500.000	10.1
Aroclor-1260-1	6.207	6.113	6.313	551.900	500.000	10.4
Aroclor-1260-2	6.396	6.302	6.502	553.970	500.000	10.8
Aroclor-1260-3	6.547	6.453	6.653	576.070	500.000	15.2
Aroclor-1260-4	7.016	6.923	7.123	560.700	500.000	12.1
Aroclor-1260-5	7.258	7.164	7.364	581.870	500.000	16.4
Decachlorobiphenyl	8.633	8.540	8.740	55.140	50.000	10.3
Tetrachloro-m-xylene	3.659	3.562	3.762	55.610	50.000	11.2