

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2747
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL04 **Date Analyzed:** 08/05/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112744.D **Time Analyzed:** 10:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.752	4.655	4.855	492.690	500.000	-1.5
Aroclor-1016-2	4.770	4.674	4.874	484.830	500.000	-3.0
Aroclor-1016-3	4.828	4.731	4.931	480.470	500.000	-3.9
Aroclor-1016-4	4.947	4.851	5.051	480.160	500.000	-4.0
Aroclor-1016-5	5.204	5.108	5.308	494.700	500.000	-1.1
Aroclor-1260-1	6.241	6.145	6.345	480.100	500.000	-4.0
Aroclor-1260-2	6.430	6.333	6.533	469.920	500.000	-6.0
Aroclor-1260-3	6.796	6.700	6.900	498.060	500.000	-0.4
Aroclor-1260-4	7.055	6.959	7.159	469.500	500.000	-6.1
Aroclor-1260-5	7.298	7.203	7.403	468.040	500.000	-6.4
Decachlorobiphenyl	8.687	8.593	8.793	45.470	50.000	-9.1
Tetrachloro-m-xylene	3.667	3.569	3.769	48.650	50.000	-2.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2747
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL04 **Date Analyzed:** 08/05/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112744.D **Time Analyzed:** 10:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.733	4.638	4.838	447.870	500.000	-10.4
Aroclor-1016-2	4.751	4.656	4.856	445.070	500.000	-11.0
Aroclor-1016-3	4.926	4.831	5.031	438.480	500.000	-12.3
Aroclor-1016-4	4.968	4.873	5.073	443.220	500.000	-11.4
Aroclor-1016-5	5.180	5.085	5.285	448.020	500.000	-10.4
Aroclor-1260-1	6.208	6.113	6.313	456.840	500.000	-8.6
Aroclor-1260-2	6.396	6.302	6.502	457.410	500.000	-8.5
Aroclor-1260-3	6.547	6.453	6.653	472.500	500.000	-5.5
Aroclor-1260-4	7.017	6.923	7.123	472.330	500.000	-5.5
Aroclor-1260-5	7.259	7.164	7.364	485.130	500.000	-3.0
Decachlorobiphenyl	8.632	8.540	8.740	45.260	50.000	-9.5
Tetrachloro-m-xylene	3.659	3.562	3.762	46.200	50.000	-7.6