

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2747
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL05 **Date Analyzed:** 08/04/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074203.D **Time Analyzed:** 09:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.808	5.712	5.912	518.450	500.000	3.7
Aroclor-1016-2	5.830	5.733	5.933	519.980	500.000	4.0
Aroclor-1016-3	5.893	5.796	5.996	527.530	500.000	5.5
Aroclor-1016-4	5.990	5.893	6.093	530.670	500.000	6.1
Aroclor-1016-5	6.282	6.186	6.386	519.430	500.000	3.9
Aroclor-1260-1	7.399	7.303	7.503	514.320	500.000	2.9
Aroclor-1260-2	7.652	7.555	7.755	493.090	500.000	-1.4
Aroclor-1260-3	8.010	7.913	8.113	500.890	500.000	0.2
Aroclor-1260-4	8.238	8.142	8.342	508.010	500.000	1.6
Aroclor-1260-5	8.563	8.468	8.668	493.510	500.000	-1.3
Decachlorobiphenyl	10.433	10.338	10.538	50.720	50.000	1.4
Tetrachloro-m-xylene	4.657	4.560	4.760	54.520	50.000	9.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2747
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL05 **Date Analyzed:** 08/04/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074203.D **Time Analyzed:** 09:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.904	4.806	5.006	514.160	500.000	2.8
Aroclor-1016-2	4.961	4.863	5.063	520.690	500.000	4.1
Aroclor-1016-3	5.081	4.983	5.183	507.050	500.000	1.4
Aroclor-1016-4	5.122	5.023	5.223	508.480	500.000	1.7
Aroclor-1016-5	5.336	5.239	5.439	526.840	500.000	5.4
Aroclor-1260-1	6.553	6.455	6.655	492.160	500.000	-1.6
Aroclor-1260-2	6.707	6.609	6.809	490.910	500.000	-1.8
Aroclor-1260-3	6.918	6.819	7.019	501.210	500.000	0.2
Aroclor-1260-4	7.177	7.079	7.279	490.060	500.000	-2.0
Aroclor-1260-5	7.416	7.318	7.518	509.190	500.000	1.8
Decachlorobiphenyl	8.823	8.726	8.926	48.830	50.000	-2.3
Tetrachloro-m-xylene	3.803	3.705	3.905	52.420	50.000	4.8