

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2747
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL06 **Date Analyzed:** 08/04/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074214.D **Time Analyzed:** 15:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.810	5.712	5.912	518.860	500.000	3.8
Aroclor-1016-2	5.832	5.733	5.933	528.460	500.000	5.7
Aroclor-1016-3	5.895	5.796	5.996	526.530	500.000	5.3
Aroclor-1016-4	5.992	5.893	6.093	529.720	500.000	5.9
Aroclor-1016-5	6.284	6.186	6.386	528.550	500.000	5.7
Aroclor-1260-1	7.401	7.303	7.503	520.430	500.000	4.1
Aroclor-1260-2	7.654	7.555	7.755	502.070	500.000	0.4
Aroclor-1260-3	8.012	7.913	8.113	509.260	500.000	1.9
Aroclor-1260-4	8.240	8.142	8.342	513.110	500.000	2.6
Aroclor-1260-5	8.565	8.468	8.668	501.760	500.000	0.4
Decachlorobiphenyl	10.435	10.338	10.538	51.610	50.000	3.2
Tetrachloro-m-xylene	4.660	4.560	4.760	52.820	50.000	5.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2747
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL06 **Date Analyzed:** 08/04/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074214.D **Time Analyzed:** 15:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.907	4.806	5.006	538.570	500.000	7.7
Aroclor-1016-2	4.965	4.863	5.063	536.500	500.000	7.3
Aroclor-1016-3	5.084	4.983	5.183	538.190	500.000	7.6
Aroclor-1016-4	5.125	5.023	5.223	541.020	500.000	8.2
Aroclor-1016-5	5.340	5.239	5.439	541.120	500.000	8.2
Aroclor-1260-1	6.555	6.455	6.655	522.820	500.000	4.6
Aroclor-1260-2	6.710	6.609	6.809	514.440	500.000	2.9
Aroclor-1260-3	6.921	6.819	7.019	531.650	500.000	6.3
Aroclor-1260-4	7.180	7.079	7.279	523.920	500.000	4.8
Aroclor-1260-5	7.419	7.318	7.518	531.100	500.000	6.2
Decachlorobiphenyl	8.825	8.726	8.926	51.460	50.000	2.9
Tetrachloro-m-xylene	3.807	3.705	3.905	56.600	50.000	13.2