

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2747
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL07 **Date Analyzed:** 08/05/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074219.D **Time Analyzed:** 08:31

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.804	5.712	5.912	539.830	500.000	8.0
Aroclor-1016-2	5.825	5.733	5.933	534.330	500.000	6.9
Aroclor-1016-3	5.888	5.796	5.996	544.330	500.000	8.9
Aroclor-1016-4	5.985	5.893	6.093	545.670	500.000	9.1
Aroclor-1016-5	6.277	6.186	6.386	541.360	500.000	8.3
Aroclor-1260-1	7.394	7.303	7.503	527.510	500.000	5.5
Aroclor-1260-2	7.647	7.555	7.755	508.540	500.000	1.7
Aroclor-1260-3	8.004	7.913	8.113	517.320	500.000	3.5
Aroclor-1260-4	8.232	8.142	8.342	520.560	500.000	4.1
Aroclor-1260-5	8.557	8.468	8.668	505.610	500.000	1.1
Decachlorobiphenyl	10.425	10.338	10.538	51.970	50.000	3.9
Tetrachloro-m-xylene	4.653	4.560	4.760	55.200	50.000	10.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2747
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL07 **Date Analyzed:** 08/05/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074219.D **Time Analyzed:** 08:31

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.902	4.806	5.006	546.950	500.000	9.4
Aroclor-1016-2	4.960	4.863	5.063	550.450	500.000	10.1
Aroclor-1016-3	5.080	4.983	5.183	537.480	500.000	7.5
Aroclor-1016-4	5.120	5.023	5.223	566.810	500.000	13.4
Aroclor-1016-5	5.334	5.239	5.439	567.310	500.000	13.5
Aroclor-1260-1	6.551	6.455	6.655	540.310	500.000	8.1
Aroclor-1260-2	6.706	6.609	6.809	521.750	500.000	4.4
Aroclor-1260-3	6.915	6.819	7.019	553.110	500.000	10.6
Aroclor-1260-4	7.176	7.079	7.279	529.980	500.000	6.0
Aroclor-1260-5	7.414	7.318	7.518	551.700	500.000	10.3
Decachlorobiphenyl	8.820	8.726	8.926	52.050	50.000	4.1
Tetrachloro-m-xylene	3.802	3.705	3.905	57.780	50.000	15.6