

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2747
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL08 **Date Analyzed:** 08/05/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074224.D **Time Analyzed:** 10:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.809	5.712	5.912	529.600	500.000	5.9
Aroclor-1016-2	5.830	5.733	5.933	533.620	500.000	6.7
Aroclor-1016-3	5.893	5.796	5.996	536.280	500.000	7.3
Aroclor-1016-4	5.990	5.893	6.093	538.680	500.000	7.7
Aroclor-1016-5	6.283	6.186	6.386	537.510	500.000	7.5
Aroclor-1260-1	7.400	7.303	7.503	529.550	500.000	5.9
Aroclor-1260-2	7.652	7.555	7.755	506.080	500.000	1.2
Aroclor-1260-3	8.010	7.913	8.113	509.930	500.000	2.0
Aroclor-1260-4	8.239	8.142	8.342	513.310	500.000	2.7
Aroclor-1260-5	8.564	8.468	8.668	500.370	500.000	0.1
Decachlorobiphenyl	10.433	10.338	10.538	51.200	50.000	2.4
Tetrachloro-m-xylene	4.658	4.560	4.760	54.490	50.000	9.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2747
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL08 **Date Analyzed:** 08/05/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074224.D **Time Analyzed:** 10:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.904	4.806	5.006	555.950	500.000	11.2
Aroclor-1016-2	4.962	4.863	5.063	559.870	500.000	12.0
Aroclor-1016-3	5.083	4.983	5.183	546.400	500.000	9.3
Aroclor-1016-4	5.123	5.023	5.223	543.720	500.000	8.7
Aroclor-1016-5	5.337	5.239	5.439	560.300	500.000	12.1
Aroclor-1260-1	6.553	6.455	6.655	547.710	500.000	9.5
Aroclor-1260-2	6.709	6.609	6.809	526.600	500.000	5.3
Aroclor-1260-3	6.918	6.819	7.019	544.840	500.000	9.0
Aroclor-1260-4	7.178	7.079	7.279	529.000	500.000	5.8
Aroclor-1260-5	7.417	7.318	7.518	531.910	500.000	6.4
Decachlorobiphenyl	8.822	8.726	8.926	51.840	50.000	3.7
Tetrachloro-m-xylene	3.805	3.705	3.905	57.110	50.000	14.2