

Surrogate Summary

SDG No.: Q2747

Client: Nobis Group

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO112413.D	PIBLK-PO112413.D	Tetrachloro-m-xyl	1	20	19.1	95		60	140
		Decachlorobiphen	1	20	20.0	100		60	140
		Tetrachloro-m-xyl	2	20	19.5	98		60	140
		Decachlorobiphen	2	20	20.4	102		60	140
I.BLK-PO112722.D	PIBLK-PO112722.D	Tetrachloro-m-xyl	1	20	19.8	99		60	140
		Decachlorobiphen	1	20	19.2	96		60	140
		Tetrachloro-m-xyl	2	20	19.0	95		60	140
		Decachlorobiphen	2	20	19.5	98		60	140
Q2747-03	OU4-TS-GRILLO-OG3-073125	Tetrachloro-m-xyl	1	20	18.0	90		44	130
		Decachlorobiphen	1	20	12.8	64		60	125
		Tetrachloro-m-xyl	2	20	17.6	88		44	130
		Decachlorobiphen	2	20	13.8	69		60	125
Q2747-05	OU4-TS-GRILLO-OG4-073125	Tetrachloro-m-xyl	1	20	19.2	96		44	130
		Decachlorobiphen	1	20	14.0	70		60	125
		Tetrachloro-m-xyl	2	20	18.6	93		44	130
		Decachlorobiphen	2	20	14.9	74		60	125
I.BLK-PO112732.D	PIBLK-PO112732.D	Tetrachloro-m-xyl	1	20	19.1	95		60	140
		Decachlorobiphen	1	20	18.3	92		60	140
		Tetrachloro-m-xyl	2	20	18.8	94		60	140
		Decachlorobiphen	2	20	19.5	98		60	140
I.BLK-PO112741.D	PIBLK-PO112741.D	Tetrachloro-m-xyl	1	20	20.8	104		60	140
		Decachlorobiphen	1	20	19.9	100		60	140
		Tetrachloro-m-xyl	2	20	19.7	99		60	140
		Decachlorobiphen	2	20	20.5	102		60	140
Q2747-01	OU4-TS-GRILLO-OG2-073125	Tetrachloro-m-xyl	1	20	17.0	85		44	130
		Decachlorobiphen	1	20	19.3	97		60	125
		Tetrachloro-m-xyl	2	20	16.6	83		44	130
		Decachlorobiphen	2	20	20.2	101		60	125
I.BLK-PO112748.D	PIBLK-PO112748.D	Tetrachloro-m-xyl	1	20	18.1	91		60	140
		Decachlorobiphen	1	20	17.8	89		60	140
		Tetrachloro-m-xyl	2	20	17.4	87		60	140
		Decachlorobiphen	2	20	18.3	91		60	140
I.BLK-PP074167.D	PIBLK-PP074167.D	Tetrachloro-m-xyl	1	20	19.1	95		60	140
		Decachlorobiphen	1	20	19.7	98		60	140
		Tetrachloro-m-xyl	2	20	17.3	86		60	140
		Decachlorobiphen	2	20	18.7	93		60	140
I.BLK-PP074206.D	PIBLK-PP074206.D	Tetrachloro-m-xyl	1	20	21.8	109		60	140
		Decachlorobiphen	1	20	21.4	107		60	140
		Tetrachloro-m-xyl	2	20	17.9	90		60	140
		Decachlorobiphen	2	20	18.3	91		60	140
PB169092BL	PB169092BL	Tetrachloro-m-xyl	1	20	22.5	113		44	130

Surrogate Summary

SDG No.: Q2747

Client: Nobis Group

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
PB169092BL	PB169092BL	Decachlorobiphen	1	20	22.7	113		60	125
		Tetrachloro-m-xyl	2	20	19.0	95		44	130
		Decachlorobiphen	2	20	19.4	97		60	125
PB169092BS	PB169092BS	Tetrachloro-m-xyl	1	20	23.9	119		44	130
		Decachlorobiphen	1	20	23.4	117		60	125
		Tetrachloro-m-xyl	2	20	20.3	101		44	130
		Decachlorobiphen	2	20	20.4	102		60	125
Q2753-01MS	289MS	Tetrachloro-m-xyl	1	20	25.1	126		44	130
		Decachlorobiphen	1	20	21.5	107		60	125
		Tetrachloro-m-xyl	2	20	20.9	105		44	130
		Decachlorobiphen	2	20	18.8	94		60	125
Q2753-01MSD	289MSD	Tetrachloro-m-xyl	1	20	25.3	126		44	130
		Decachlorobiphen	1	20	20.7	104		60	125
		Tetrachloro-m-xyl	2	20	21.6	108		44	130
		Decachlorobiphen	2	20	18.4	92		60	125
I.BLK-PP074217.D	PIBLK-PP074217.D	Tetrachloro-m-xyl	1	20	21.9	110		60	140
		Decachlorobiphen	1	20	21.8	109		60	140
		Tetrachloro-m-xyl	2	20	18.3	91		60	140
		Decachlorobiphen	2	20	18.9	95		60	140
I.BLK-PP074220.D	PIBLK-PP074220.D	Tetrachloro-m-xyl	1	20	22.4	112		60	140
		Decachlorobiphen	1	20	22.1	110		60	140
		Tetrachloro-m-xyl	2	20	19.1	96		60	140
		Decachlorobiphen	2	20	19.7	98		60	140
Q2747-07	OU4-TS-GRILLO-TSCP01-0731	Tetrachloro-m-xyl	1	20	17.0	85		44	130
		Decachlorobiphen	1	20	21.1	106		60	125
		Tetrachloro-m-xyl	2	20	13.6	68		44	130
		Decachlorobiphen	2	20	18.3	92		60	125
Q2747-09	OU4-TS-GRILLO-TSCP02-0731	Tetrachloro-m-xyl	1	20	16.8	84		44	130
		Decachlorobiphen	1	20	17.5	88		60	125
		Tetrachloro-m-xyl	2	20	12.1	61		44	130
		Decachlorobiphen	2	20	14.5	72		60	125
I.BLK-PP074227.D	PIBLK-PP074227.D	Tetrachloro-m-xyl	1	20	22.4	112		60	140
		Decachlorobiphen	1	20	21.8	109		60	140
		Tetrachloro-m-xyl	2	20	19.0	95		60	140
		Decachlorobiphen	2	20	19.3	96		60	140