



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

**SDG No.:** Q2747

**Analytical Method:** 8082A

**Client:** Nobis Group

**Datafile :** PP074209.D

| Lab Sample ID            | Parameter | Spike | Result | Units | Rec | RPD | Qual | RPD  | Low | Limits | RPD |
|--------------------------|-----------|-------|--------|-------|-----|-----|------|------|-----|--------|-----|
|                          |           |       |        |       |     |     |      | Qual |     | High   |     |
| PB169092BS<br>(Column 1) | AR1016    | 166.5 | 178    | ug/kg | 107 |     |      |      | 47  | 134    |     |
|                          | AR1260    | 166.5 | 164    | ug/kg | 98  |     |      |      | 53  | 140    |     |
| PB169092BS<br>(Column 2) | AR1016    | 166.5 | 171    | ug/kg | 103 |     |      |      | 47  | 134    |     |
|                          | AR1260    | 166.5 | 162    | ug/kg | 97  |     |      |      | 53  | 140    |     |