

Surrogate Summary

SDG No.: Q2747

Client: Nobis Group

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD089685.D	PIBLK-PD089685.D	Decachlorobiphen	1	20	20.9	104		30	135
		Tetrachloro-m-xyl	1	20	19.6	98		44	124
		Decachlorobiphen	2	20	20.9	104		30	135
		Tetrachloro-m-xyl	2	20	20.4	102		44	124
I.BLK-PD089773.D	PIBLK-PD089773.D	Decachlorobiphen	1	20	21.1	106		30	135
		Tetrachloro-m-xyl	1	20	20.2	101		44	124
		Decachlorobiphen	2	20	24.1	120		30	135
		Tetrachloro-m-xyl	2	20	24.1	120		44	124
Q2747-01	OU4-TS-GRILLO-OG2-073125	Decachlorobiphen	1	20	12.0	60		55	130
		Tetrachloro-m-xyl	1	20	10.8	54		42	129
		Decachlorobiphen	2	20	13.7	68		55	130
		Tetrachloro-m-xyl	2	20	12.9	64		42	129
Q2747-03	OU4-TS-GRILLO-OG3-073125	Decachlorobiphen	1	20	15.8	79		55	130
		Tetrachloro-m-xyl	1	20	12.0	60		42	129
		Decachlorobiphen	2	20	17.2	86		55	130
		Tetrachloro-m-xyl	2	20	14.9	75		42	129
Q2747-05	OU4-TS-GRILLO-OG4-073125	Decachlorobiphen	1	20	11.6	58		55	130
		Tetrachloro-m-xyl	1	20	13.0	65		42	129
		Decachlorobiphen	2	20	12.7	63		55	130
		Tetrachloro-m-xyl	2	20	15.5	77		42	129
Q2747-07	OU4-TS-GRILLO-TSCP01-0731	Decachlorobiphen	1	20	13.3	67		55	130
		Tetrachloro-m-xyl	1	20	9.18	46		42	129
		Decachlorobiphen	2	20	14.9	75		55	130
		Tetrachloro-m-xyl	2	20	11.0	55		42	129
I.BLK-PD089784.D	PIBLK-PD089784.D	Decachlorobiphen	1	20	20.7	104		30	135
		Tetrachloro-m-xyl	1	20	20.1	101		44	124
		Decachlorobiphen	2	20	22.8	114		30	135
		Tetrachloro-m-xyl	2	20	24.4	122		44	124
I.BLK-PD089787.D	PIBLK-PD089787.D	Decachlorobiphen	1	20	21.5	107		30	135
		Tetrachloro-m-xyl	1	20	21.1	105		44	124
		Decachlorobiphen	2	20	23.9	120		30	135
		Tetrachloro-m-xyl	2	20	25.7	128	*	44	124
Q2747-09	OU4-TS-GRILLO-TSCP02-0731	Decachlorobiphen	1	20	10.5	52	*	55	130
		Tetrachloro-m-xyl	1	20	8.30	42		42	129
		Decachlorobiphen	2	20	12.3	61		55	130
		Tetrachloro-m-xyl	2	20	9.85	49		42	129
I.BLK-PD089792.D	PIBLK-PD089792.D	Decachlorobiphen	1	20	21.8	109		30	135
		Tetrachloro-m-xyl	1	20	21.2	106		44	124
		Decachlorobiphen	2	20	24.3	122		30	135
		Tetrachloro-m-xyl	2	20	25.5	128	*	44	124
I.BLK-PL096593.D	PIBLK-PL096593.D	Decachlorobiphen	1	20	19.2	96		30	135

Surrogate Summary

SDG No.: Q2747

Client: Nobis Group

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PL096593.D	PIBLK-PL096593.D	Tetrachloro-m-xyl	1	20	18.6	93		44	124
		Decachlorobiphen	2	20	19.1	95		30	135
I.BLK-PL096687.D	PIBLK-PL096687.D	Tetrachloro-m-xyl	2	20	18.3	92		44	124
		Decachlorobiphen	1	20	24.6	123		30	135
		Tetrachloro-m-xyl	1	20	23.1	116		44	124
		Decachlorobiphen	2	20	24.3	121		30	135
		Tetrachloro-m-xyl	2	20	23.8	119		44	124
PB169093BL	PB169093BL	Decachlorobiphen	1	20	20.3	102		55	130
		Tetrachloro-m-xyl	1	20	18.8	94		42	129
		Decachlorobiphen	2	20	21.4	107		55	130
		Tetrachloro-m-xyl	2	20	19.2	96		42	129
PB169093BS	PB169093BS	Decachlorobiphen	1	20	21.9	110		55	130
		Tetrachloro-m-xyl	1	20	20.4	102		42	129
		Decachlorobiphen	2	20	22.7	114		55	130
		Tetrachloro-m-xyl	2	20	20.9	104		42	129
Q2734-01MS	VNJ-253MS	Decachlorobiphen	1	20	20.9	104		55	130
		Tetrachloro-m-xyl	1	20	17.3	86		42	129
		Decachlorobiphen	2	20	24.5	123		55	130
Q2734-01MSD	VNJ-253MSD	Tetrachloro-m-xyl	2	20	21.4	107		42	129
		Decachlorobiphen	1	20	20.1	101		55	130
		Tetrachloro-m-xyl	1	20	16.6	83		42	129
		Decachlorobiphen	2	20	21.5	108		55	130
		Tetrachloro-m-xyl	2	20	21.2	106		42	129
I.BLK-PL096698.D	PIBLK-PL096698.D	Decachlorobiphen	1	20	23.0	115		30	135
		Tetrachloro-m-xyl	1	20	21.8	109		44	124
		Decachlorobiphen	2	20	20.1	100		30	135
		Tetrachloro-m-xyl	2	20	22.3	112		44	124