

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF071725

Review By	Rahul	Review On	7/18/2025 11:02:11 AM
Supervise By	Jagrut	Supervise On	7/18/2025 1:30:54 PM
SubDirectory	BF071725	HP Acquire Method	BNA_F
		HP Processing Method	bf071725

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12674,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF143138.D	17 Jul 2025 09:58		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF143139.D	17 Jul 2025 10:28	A Fresh Calibration is required.	RC/JU	Not Ok
3	SSTDICC2.5	SSTDICC2.5	BF143140.D	17 Jul 2025 11:04		RC/JU	Ok
4	SSTDICC005	SSTDICC005	BF143141.D	17 Jul 2025 11:34		RC/JU	Ok,M
5	SSTDICC010	SSTDICC010	BF143142.D	17 Jul 2025 12:04		RC/JU	Ok,M
6	SSTDICC020	SSTDICC020	BF143143.D	17 Jul 2025 12:34		RC/JU	Ok,M
7	SSTDICCC040	SSTDICCC040	BF143144.D	17 Jul 2025 13:03	Compound #32,54,65 Kept on LR	RC/JU	Ok,M
8	SSTDICC050	SSTDICC050	BF143145.D	17 Jul 2025 13:33		RC/JU	Ok,M
9	SSTDICC060	SSTDICC060	BF143146.D	17 Jul 2025 14:04		RC/JU	Ok,M
10	SSTDICC080	SSTDICC080	BF143147.D	17 Jul 2025 14:34	Compound#77 removed from 80 ppm	RC/JU	Ok,M
11	SSTDICV040	ICVBF071725	BF143148.D	17 Jul 2025 15:06		RC/JU	Ok,M
12	PB168813BL	PB168813BL	BF143149.D	17 Jul 2025 15:36		RC/JU	Ok
13	Q2126-03	MDL-SOIL-03-QT2-202	BF143150.D	17 Jul 2025 16:06	MDL-SOIL 4 ppm	RC/JU	Ok,M
14	Q2126-03	MDL-SOIL-03-QT2-202	BF143151.D	17 Jul 2025 16:36	MDL-SOIL 8 ppm	RC/JU	Ok,M
15	Q2126-09	MDL-WATER-03-QT2-2	BF143152.D	17 Jul 2025 17:07	MDL-WATER 4 ppm	RC/JU	Ok,M
16	Q2126-09	MDL-WATER-03-QT2-2	BF143153.D	17 Jul 2025 17:37	MDL-WATER 8 ppm	RC/JU	Ok,M
17	PB168816BL	PB168816BL	BF143154.D	17 Jul 2025 18:07		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF071725

Review By	Rahul	Review On	7/18/2025 11:02:11 AM		
Supervise By	Jagrut	Supervise On	7/18/2025 1:30:54 PM		
SubDirectory	BF071725	HP Acquire Method	BNA_F	HP Processing Method	bf071725

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12674,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	PB167904BL	PB167904BL	BF143155.D	17 Jul 2025 18:37		RC/JU	Ok
19	PB167904BS	PB167904BS	BF143156.D	17 Jul 2025 19:07		RC/JU	Ok,M
20	SSTDCCC040	SSTDCCC040EC	BF143157.D	17 Jul 2025 19:37		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF080525

Review By	Rahul	Review On	8/6/2025 10:17:44 AM
Supervise By	Jagrut	Supervise On	8/6/2025 10:56:56 AM
SubDirectory	BF080525	HP Acquire Method	BNA_F
		HP Processing Method	bf071725

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S13169,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF143304.D	05 Aug 2025 09:02		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF143305.D	05 Aug 2025 10:00		RC/JU	Ok
3	PB169095BL	PB169095BL	BF143306.D	05 Aug 2025 10:29		RC/JU	Ok
4	PB169095BS	PB169095BS	BF143307.D	05 Aug 2025 10:58		RC/JU	Ok,M
5	Q2747-01	OU4-TS-GRILLO-OG2	BF143308.D	05 Aug 2025 11:34		RC/JU	Ok,M
6	Q2747-03	OU4-TS-GRILLO-OG3	BF143309.D	05 Aug 2025 12:03		RC/JU	Ok,M
7	Q2747-05	OU4-TS-GRILLO-OG4	BF143310.D	05 Aug 2025 12:32		RC/JU	Ok
8	Q2747-07	OU4-TS-GRILLO-TSCF	BF143311.D	05 Aug 2025 13:01		RC/JU	Ok,M
9	Q2747-09	OU4-TS-GRILLO-TSCF	BF143312.D	05 Aug 2025 13:30		RC/JU	Ok,M
10	Q2734-01MS	VNJ-253MS	BF143313.D	05 Aug 2025 13:59		RC/JU	Ok,M
11	Q2734-01MSD	VNJ-253MSD	BF143314.D	05 Aug 2025 14:28		RC/JU	Ok,M
12	Q2759-01	SOLID-DRUMS	BF143315.D	05 Aug 2025 14:57	Need 5X.	RC/JU	Dilution
13	Q2758-01	WASTE	BF143316.D	05 Aug 2025 15:26		RC/JU	Ok,M
14	Q2760-01	NB-07-08042025	BF143317.D	05 Aug 2025 15:55		RC/JU	Ok
15	Q2759-01DL	SOLID-DRUMSDL	BF143318.D	05 Aug 2025 16:24		RC/JU	Ok
16	SSTDCCC040	SSTDCCC040EC	BF143319.D	05 Aug 2025 16:54		RC/JU	Ok,M

M : Manual Integration