

Surrogate Summary

SW-846

SDG No.: Q2747

Client: Nobis Group

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB169095BL	PB169095BL	Nitrobenzene-d5	100	85.1	85		37	122
		2-Fluorobiphenyl	100	82.1	82		44	115
		Terphenyl-d14	100	82.4	82		54	127
PB169095BS	PB169095BS	Nitrobenzene-d5	100	82.0	82		37	122
		2-Fluorobiphenyl	100	79.4	79		44	115
		Terphenyl-d14	100	80.3	80		54	127
Q2734-01MS	VNJ-253MS	Nitrobenzene-d5	100	56.4	56		37	122
		2-Fluorobiphenyl	100	59.0	59		44	115
		Terphenyl-d14	100	47.0	47	*	54	127
Q2734-01MSD	VNJ-253MSD	Nitrobenzene-d5	100	60.6	61		37	122
		2-Fluorobiphenyl	100	64.5	64		44	115
		Terphenyl-d14	100	50.7	51	*	54	127
Q2747-01	OU4-TS-GRILLO-OG2-073125	Nitrobenzene-d5	100	57.1	57		37	122
		2-Fluorobiphenyl	100	55.7	56		44	115
		Terphenyl-d14	100	41.8	42	*	54	127
Q2747-03	OU4-TS-GRILLO-OG3-073125	Nitrobenzene-d5	100	55.7	56		37	122
		2-Fluorobiphenyl	100	53.4	53		44	115
		Terphenyl-d14	100	38.0	38	*	54	127
Q2747-05	OU4-TS-GRILLO-OG4-073125	Nitrobenzene-d5	100	59.0	59		37	122
		2-Fluorobiphenyl	100	59.0	59		44	115
		Terphenyl-d14	100	39.4	39	*	54	127
Q2747-07	OU4-TS-GRILLO-TSCP01-073125	Nitrobenzene-d5	100	50.6	51		37	122
		2-Fluorobiphenyl	100	50.3	50		44	115
		Terphenyl-d14	100	40.3	40	*	54	127
Q2747-09	OU4-TS-GRILLO-TSCP02-073125	Nitrobenzene-d5	100	55.5	56		37	122
		2-Fluorobiphenyl	100	57.4	57		44	115
		Terphenyl-d14	100	42.2	42	*	54	127