

Continuing Calibration Report for SequenceID : FE081225AL

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FE055243.D

|                   |              |         |        |        |            |            |        |
|-------------------|--------------|---------|--------|--------|------------|------------|--------|
| Aliphatic C9-C12  | 6253802.000  | 60.000  | 3.323  | 6.959  | 104230.033 | 116782.597 | 10.749 |
| Aliphatic C12-C16 | 4578727.000  | 40.000  | 6.960  | 10.413 | 114468.175 | 125227.070 | 8.592  |
| Aliphatic C16-C21 | 6890938.000  | 60.000  | 10.414 | 13.793 | 114848.967 | 126404.887 | 9.142  |
| Aliphatic C21-C28 | 9492701.000  | 80.000  | 13.794 | 17.466 | 118658.763 | 120505.719 | 1.533  |
| Aliphatic C28-C40 | 14343816.000 | 120.000 | 17.467 | 22.498 | 119531.800 | 113987.195 | -4.864 |
| Aliphatic EPH     | 41559984.000 | 360.000 | 3.323  | 22.498 | 115444.400 | 119220.147 | 3.167  |