

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: SCIA01
 Lab Code: ACE SDG NO.: Q2757
 Lab File ID: VW032004.D Date Analyzed: 08/05/2025
 Instrument ID: MSVOA_W Time Analyzed: 09:05
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	203098	7.95	376001	8.85	340179	11.63
UPPER LIMIT	406196	8.453	752002	9.349	680358	12.129
LOWER LIMIT	101549	7.453	188001	8.349	170090	11.129
EPA SAMPLE NO.						
VOC	142857	7.96	290928	8.86	269684	11.64
VW0805SBL01	172047	7.96	368948	8.85	349230	11.63
VW0805SBS01	208310	7.95	391172	8.85	361011	11.63
VW0805SBSD01	231820	7.96	413781	8.85	373116	11.63

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.