

Initial and Continuing Calibration Blank Summary

Client: Sciacca General Contractors, LLC

SDG No.: Q2758

Project: 92 Clifford Street, Newark

RunNo.: LB136704

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Reactive Cyanide	mg/L	0.001	0.0025	J	0.00096	0.005	08/05/2025
Sample ID: CCB1 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	08/05/2025
Sample ID: CCB2 Reactive Cyanide	mg/L	0.0012	0.0025	J	0.00096	0.005	08/05/2025
Sample ID: CCB3 Reactive Cyanide	mg/L	0.00096	0.0025	J	0.00096	0.005	08/05/2025

Initial and Continuing Calibration Blank Summary

Client: Sciacca General Contractors, LLC

SDG No.: Q2758

Project: 92 Clifford Street, Newark

RunNo.: LB136704

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
---------	-------	--------	----------------------	--------------	-----	-----	------------------