



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Sciacca General Contractors, LLC

SDG No.: Q2758

Contract: SCIA01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB51	Mercury	0.076	+/-0.2	U	0.20	CV	08/04/2025	14:02	LB136689

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Sciaccia General Contractors, LLC

SDG No.: Q2758

Contract: SCIA01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB80	Mercury	0.076	+/-0.2	U	0.20	CV	08/04/2025	14:07	LB136689
CCB81	Mercury	0.076	+/-0.2	U	0.20	CV	08/04/2025	15:28	LB136689
CCB82	Mercury	0.076	+/-0.2	U	0.20	CV	08/04/2025	15:51	LB136689

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Sciacca General Contractors, LLC

SDG No.: Q2758

Contract: SCIA01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	08/04/2025	11:52	LB136697
	Antimony	6.76	+/-25	U	50.0	P	08/04/2025	11:52	LB136697
	Arsenic	5.12	+/-10	U	20.0	P	08/04/2025	11:52	LB136697
	Barium	14.6	+/-50	U	100	P	08/04/2025	11:52	LB136697
	Beryllium	0.56	+/-3	U	6.00	P	08/04/2025	11:52	LB136697
	Cadmium	0.50	+/-3	U	6.00	P	08/04/2025	11:52	LB136697
	Calcium	234	+/-1000	U	2000	P	08/04/2025	11:52	LB136697
	Chromium	2.12	+/-5	U	10.0	P	08/04/2025	11:52	LB136697
	Cobalt	2.26	+/-15	U	30.0	P	08/04/2025	11:52	LB136697
	Copper	4.60	+/-10	U	20.0	P	08/04/2025	11:52	LB136697
	Iron	23.4	+/-50	U	100	P	08/04/2025	11:52	LB136697
	Lead	2.30	+/-6	U	12.0	P	08/04/2025	11:52	LB136697
	Magnesium	244	+/-1000	U	2000	P	08/04/2025	11:52	LB136697
	Manganese	5.94	+/-10	U	20.0	P	08/04/2025	11:52	LB136697
	Nickel	3.06	+/-20	U	40.0	P	08/04/2025	11:52	LB136697
	Potassium	918	+/-1000	U	2000	P	08/04/2025	11:52	LB136697
	Selenium	9.64	+/-10	U	20.0	P	08/04/2025	11:52	LB136697
	Silver	1.62	+/-5	U	10.0	P	08/04/2025	11:52	LB136697
	Sodium	868	+/-1000	U	2000	P	08/04/2025	11:52	LB136697
	Thallium	4.38	+/-20	U	40.0	P	08/04/2025	11:52	LB136697
	Vanadium	6.26	+/-20	U	40.0	P	08/04/2025	11:52	LB136697
	Zinc	16.7	+/-20	U	40.0	P	08/04/2025	11:52	LB136697

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Sciacca General Contractors, LLC

SDG No.: Q2758

Contract: SCIA01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	08/04/2025	12:29	LB136697
	Antimony	6.76	+/-25	U	50.0	P	08/04/2025	12:29	LB136697
	Arsenic	5.12	+/-10	U	20.0	P	08/04/2025	12:29	LB136697
	Barium	14.6	+/-50	U	100	P	08/04/2025	12:29	LB136697
	Beryllium	0.56	+/-3	U	6.00	P	08/04/2025	12:29	LB136697
	Cadmium	0.50	+/-3	U	6.00	P	08/04/2025	12:29	LB136697
	Calcium	234	+/-1000	U	2000	P	08/04/2025	12:29	LB136697
	Chromium	2.12	+/-5	U	10.0	P	08/04/2025	12:29	LB136697
	Cobalt	2.26	+/-15	U	30.0	P	08/04/2025	12:29	LB136697
	Copper	4.60	+/-10	U	20.0	P	08/04/2025	12:29	LB136697
	Iron	23.4	+/-50	U	100	P	08/04/2025	12:29	LB136697
	Lead	2.30	+/-6	U	12.0	P	08/04/2025	12:29	LB136697
	Magnesium	244	+/-1000	U	2000	P	08/04/2025	12:29	LB136697
	Manganese	5.94	+/-10	U	20.0	P	08/04/2025	12:29	LB136697
	Nickel	3.06	+/-20	U	40.0	P	08/04/2025	12:29	LB136697
	Potassium	918	+/-1000	U	2000	P	08/04/2025	12:29	LB136697
	Selenium	9.64	+/-10	U	20.0	P	08/04/2025	12:29	LB136697
	Silver	1.62	+/-5	U	10.0	P	08/04/2025	12:29	LB136697
	Sodium	868	+/-1000	U	2000	P	08/04/2025	12:29	LB136697
	Thallium	4.38	+/-20	U	40.0	P	08/04/2025	12:29	LB136697
	Vanadium	6.26	+/-20	U	40.0	P	08/04/2025	12:29	LB136697
	Zinc	16.7	+/-20	U	40.0	P	08/04/2025	12:29	LB136697
CCB02	Aluminum	12.9	+/-50	J	100	P	08/04/2025	13:41	LB136697
	Antimony	6.76	+/-25	U	50.0	P	08/04/2025	13:41	LB136697
	Arsenic	5.12	+/-10	U	20.0	P	08/04/2025	13:41	LB136697
	Barium	14.6	+/-50	U	100	P	08/04/2025	13:41	LB136697
	Beryllium	0.56	+/-3	U	6.00	P	08/04/2025	13:41	LB136697
	Cadmium	0.50	+/-3	U	6.00	P	08/04/2025	13:41	LB136697
	Calcium	234	+/-1000	U	2000	P	08/04/2025	13:41	LB136697
	Chromium	2.12	+/-5	U	10.0	P	08/04/2025	13:41	LB136697
	Cobalt	2.26	+/-15	U	30.0	P	08/04/2025	13:41	LB136697
	Copper	4.60	+/-10	U	20.0	P	08/04/2025	13:41	LB136697
	Iron	23.4	+/-50	U	100	P	08/04/2025	13:41	LB136697
	Lead	2.30	+/-6	U	12.0	P	08/04/2025	13:41	LB136697
	Magnesium	244	+/-1000	U	2000	P	08/04/2025	13:41	LB136697
	Manganese	5.94	+/-10	U	20.0	P	08/04/2025	13:41	LB136697
	Nickel	3.06	+/-20	U	40.0	P	08/04/2025	13:41	LB136697
	Potassium	918	+/-1000	U	2000	P	08/04/2025	13:41	LB136697
	Selenium	9.64	+/-10	U	20.0	P	08/04/2025	13:41	LB136697

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Sciacca General Contractors, LLC

SDG No.: Q2758

Contract: SCIA01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	08/04/2025	13:41	LB136697
	Sodium	868	+/-1000	U	2000	P	08/04/2025	13:41	LB136697
	Thallium	4.38	+/-20	U	40.0	P	08/04/2025	13:41	LB136697
	Vanadium	6.26	+/-20	U	40.0	P	08/04/2025	13:41	LB136697
	Zinc	16.7	+/-20	U	40.0	P	08/04/2025	13:41	LB136697
CCB03	Aluminum	11.3	+/-50	U	100	P	08/04/2025	14:54	LB136697
	Antimony	6.76	+/-25	U	50.0	P	08/04/2025	14:54	LB136697
	Arsenic	5.12	+/-10	U	20.0	P	08/04/2025	14:54	LB136697
	Barium	14.6	+/-50	U	100	P	08/04/2025	14:54	LB136697
	Beryllium	0.56	+/-3	U	6.00	P	08/04/2025	14:54	LB136697
	Cadmium	0.50	+/-3	U	6.00	P	08/04/2025	14:54	LB136697
	Calcium	234	+/-1000	U	2000	P	08/04/2025	14:54	LB136697
	Chromium	2.12	+/-5	U	10.0	P	08/04/2025	14:54	LB136697
	Cobalt	2.26	+/-15	U	30.0	P	08/04/2025	14:54	LB136697
	Copper	4.60	+/-10	U	20.0	P	08/04/2025	14:54	LB136697
	Iron	23.4	+/-50	U	100	P	08/04/2025	14:54	LB136697
	Lead	2.30	+/-6	U	12.0	P	08/04/2025	14:54	LB136697
	Magnesium	244	+/-1000	U	2000	P	08/04/2025	14:54	LB136697
	Manganese	5.94	+/-10	U	20.0	P	08/04/2025	14:54	LB136697
	Nickel	3.06	+/-20	U	40.0	P	08/04/2025	14:54	LB136697
	Potassium	918	+/-1000	U	2000	P	08/04/2025	14:54	LB136697
	Selenium	9.64	+/-10	U	20.0	P	08/04/2025	14:54	LB136697
	Silver	1.62	+/-5	U	10.0	P	08/04/2025	14:54	LB136697
	Sodium	868	+/-1000	U	2000	P	08/04/2025	14:54	LB136697
	Thallium	4.38	+/-20	U	40.0	P	08/04/2025	14:54	LB136697
	Vanadium	6.26	+/-20	U	40.0	P	08/04/2025	14:54	LB136697
	Zinc	16.7	+/-20	U	40.0	P	08/04/2025	14:54	LB136697
CCB04	Aluminum	22.0	+/-50	J	100	P	08/04/2025	15:53	LB136697
	Antimony	6.76	+/-25	U	50.0	P	08/04/2025	15:53	LB136697
	Arsenic	5.12	+/-10	U	20.0	P	08/04/2025	15:53	LB136697
	Barium	14.6	+/-50	U	100	P	08/04/2025	15:53	LB136697
	Beryllium	0.56	+/-3	U	6.00	P	08/04/2025	15:53	LB136697
	Cadmium	0.50	+/-3	U	6.00	P	08/04/2025	15:53	LB136697
	Calcium	234	+/-1000	U	2000	P	08/04/2025	15:53	LB136697
	Chromium	2.12	+/-5	U	10.0	P	08/04/2025	15:53	LB136697
	Cobalt	2.26	+/-15	U	30.0	P	08/04/2025	15:53	LB136697
	Copper	4.60	+/-10	U	20.0	P	08/04/2025	15:53	LB136697
	Iron	23.4	+/-50	U	100	P	08/04/2025	15:53	LB136697
	Lead	2.30	+/-6	U	12.0	P	08/04/2025	15:53	LB136697

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Sciacca General Contractors, LLC

SDG No.: Q2758

Contract: SCIA01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	08/04/2025	15:53	LB136697
	Manganese	5.94	+/-10	U	20.0	P	08/04/2025	15:53	LB136697
	Nickel	3.06	+/-20	U	40.0	P	08/04/2025	15:53	LB136697
	Potassium	918	+/-1000	U	2000	P	08/04/2025	15:53	LB136697
	Selenium	9.64	+/-10	U	20.0	P	08/04/2025	15:53	LB136697
	Silver	1.62	+/-5	U	10.0	P	08/04/2025	15:53	LB136697
	Sodium	868	+/-1000	U	2000	P	08/04/2025	15:53	LB136697
	Thallium	4.38	+/-20	U	40.0	P	08/04/2025	15:53	LB136697
	Vanadium	6.26	+/-20	U	40.0	P	08/04/2025	15:53	LB136697
	Zinc	16.7	+/-20	U	40.0	P	08/04/2025	15:53	LB136697
CCB05	Aluminum	11.3	+/-50	U	100	P	08/04/2025	16:55	LB136697
	Antimony	6.76	+/-25	U	50.0	P	08/04/2025	16:55	LB136697
	Arsenic	5.12	+/-10	U	20.0	P	08/04/2025	16:55	LB136697
	Barium	14.6	+/-50	U	100	P	08/04/2025	16:55	LB136697
	Beryllium	0.56	+/-3	U	6.00	P	08/04/2025	16:55	LB136697
	Cadmium	0.50	+/-3	U	6.00	P	08/04/2025	16:55	LB136697
	Calcium	234	+/-1000	U	2000	P	08/04/2025	16:55	LB136697
	Chromium	2.12	+/-5	U	10.0	P	08/04/2025	16:55	LB136697
	Cobalt	2.26	+/-15	U	30.0	P	08/04/2025	16:55	LB136697
	Copper	4.60	+/-10	U	20.0	P	08/04/2025	16:55	LB136697
	Iron	23.4	+/-50	U	100	P	08/04/2025	16:55	LB136697
	Lead	2.30	+/-6	U	12.0	P	08/04/2025	16:55	LB136697
	Magnesium	244	+/-1000	U	2000	P	08/04/2025	16:55	LB136697
	Manganese	5.94	+/-10	U	20.0	P	08/04/2025	16:55	LB136697
	Nickel	3.06	+/-20	U	40.0	P	08/04/2025	16:55	LB136697
	Potassium	918	+/-1000	U	2000	P	08/04/2025	16:55	LB136697
	Selenium	9.64	+/-10	U	20.0	P	08/04/2025	16:55	LB136697
	Silver	1.62	+/-5	U	10.0	P	08/04/2025	16:55	LB136697
	Sodium	868	+/-1000	U	2000	P	08/04/2025	16:55	LB136697
	Thallium	4.38	+/-20	U	40.0	P	08/04/2025	16:55	LB136697
	Vanadium	6.26	+/-20	U	40.0	P	08/04/2025	16:55	LB136697
	Zinc	16.7	+/-20	U	40.0	P	08/04/2025	16:55	LB136697
CCB06	Aluminum	11.3	+/-50	U	100	P	08/04/2025	17:45	LB136697
	Antimony	6.76	+/-25	U	50.0	P	08/04/2025	17:45	LB136697
	Arsenic	5.12	+/-10	U	20.0	P	08/04/2025	17:45	LB136697
	Barium	14.6	+/-50	U	100	P	08/04/2025	17:45	LB136697
	Beryllium	0.56	+/-3	U	6.00	P	08/04/2025	17:45	LB136697
	Cadmium	0.50	+/-3	U	6.00	P	08/04/2025	17:45	LB136697
	Calcium	234	+/-1000	U	2000	P	08/04/2025	17:45	LB136697

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Sciacca General Contractors, LLC

SDG No.: Q2758

Contract: SCIA01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.12	+/-5	U	10.0	P	08/04/2025	17:45	LB136697
	Cobalt	2.26	+/-15	U	30.0	P	08/04/2025	17:45	LB136697
	Copper	4.60	+/-10	U	20.0	P	08/04/2025	17:45	LB136697
	Iron	23.4	+/-50	U	100	P	08/04/2025	17:45	LB136697
	Lead	2.30	+/-6	U	12.0	P	08/04/2025	17:45	LB136697
	Magnesium	244	+/-1000	U	2000	P	08/04/2025	17:45	LB136697
	Manganese	5.94	+/-10	U	20.0	P	08/04/2025	17:45	LB136697
	Nickel	3.06	+/-20	U	40.0	P	08/04/2025	17:45	LB136697
	Potassium	918	+/-1000	U	2000	P	08/04/2025	17:45	LB136697
	Selenium	9.64	+/-10	U	20.0	P	08/04/2025	17:45	LB136697
	Silver	1.62	+/-5	U	10.0	P	08/04/2025	17:45	LB136697
	Sodium	868	+/-1000	U	2000	P	08/04/2025	17:45	LB136697
	Thallium	4.38	+/-20	U	40.0	P	08/04/2025	17:45	LB136697
	Vanadium	6.26	+/-20	U	40.0	P	08/04/2025	17:45	LB136697
	Zinc	16.7	+/-20	U	40.0	P	08/04/2025	17:45	LB136697
CCB07	Aluminum	11.3	+/-50	U	100	P	08/04/2025	18:02	LB136697
	Antimony	6.76	+/-25	U	50.0	P	08/04/2025	18:02	LB136697
	Arsenic	5.12	+/-10	U	20.0	P	08/04/2025	18:02	LB136697
	Barium	14.6	+/-50	U	100	P	08/04/2025	18:02	LB136697
	Beryllium	0.56	+/-3	U	6.00	P	08/04/2025	18:02	LB136697
	Cadmium	0.50	+/-3	U	6.00	P	08/04/2025	18:02	LB136697
	Calcium	234	+/-1000	U	2000	P	08/04/2025	18:02	LB136697
	Chromium	2.12	+/-5	U	10.0	P	08/04/2025	18:02	LB136697
	Cobalt	2.26	+/-15	U	30.0	P	08/04/2025	18:02	LB136697
	Copper	4.60	+/-10	U	20.0	P	08/04/2025	18:02	LB136697
	Iron	23.4	+/-50	U	100	P	08/04/2025	18:02	LB136697
	Lead	2.30	+/-6	U	12.0	P	08/04/2025	18:02	LB136697
	Magnesium	244	+/-1000	U	2000	P	08/04/2025	18:02	LB136697
	Manganese	5.94	+/-10	U	20.0	P	08/04/2025	18:02	LB136697
	Nickel	3.06	+/-20	U	40.0	P	08/04/2025	18:02	LB136697
	Potassium	918	+/-1000	U	2000	P	08/04/2025	18:02	LB136697
	Selenium	9.64	+/-10	U	20.0	P	08/04/2025	18:02	LB136697
	Silver	1.62	+/-5	U	10.0	P	08/04/2025	18:02	LB136697
	Sodium	868	+/-1000	U	2000	P	08/04/2025	18:02	LB136697
	Thallium	4.38	+/-20	U	40.0	P	08/04/2025	18:02	LB136697
	Vanadium	6.26	+/-20	U	40.0	P	08/04/2025	18:02	LB136697
	Zinc	16.7	+/-20	U	40.0	P	08/04/2025	18:02	LB136697