



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Sciacca General Contractors, LLC

SDG No.: Q2758

Contract: SCIA01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	18.9	20.0	94	65 - 135	P	08/07/2025	13:29	LB136746
	Barium	107	100	107	65 - 135	P	08/07/2025	13:29	LB136746
	Cadmium	6.46	6.0	108	65 - 135	P	08/07/2025	13:29	LB136746
	Chromium	10.5	10.0	105	65 - 135	P	08/07/2025	13:29	LB136746
	Lead	13.6	12.0	113	65 - 135	P	08/07/2025	13:29	LB136746
	Selenium	23.4	20.0	117	65 - 135	P	08/07/2025	13:29	LB136746
	Silver	10.1	10.0	101	65 - 135	P	08/07/2025	13:29	LB136746
CRA	Mercury	0.17	0.2	84	70 - 130	CV	08/07/2025	14:43	LB136741