



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Sciacca General Contractors, LLC

SDG No.: Q2758

Contract: SCIA01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB03	Mercury	0.076	+/-0.2	U	0.20	CV	08/07/2025	14:34	LB136741

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Sciaccia General Contractors, LLC

SDG No.: Q2758

Contract: SCIA01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB05	Mercury	0.076	+/-0.2	U	0.20	CV	08/07/2025	14:41	LB136741
CCB06	Mercury	0.076	+/-0.2	U	0.20	CV	08/07/2025	15:11	LB136741
CCB07	Mercury	0.076	+/-0.2	U	0.20	CV	08/07/2025	15:36	LB136741

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Sciacca General Contractors, LLC

SDG No.: Q2758

Contract: SCIA01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	08/07/2025	13:25	LB136746
	Barium	14.6	+/-50	U	100	P	08/07/2025	13:25	LB136746
	Cadmium	0.50	+/-3	U	6.00	P	08/07/2025	13:25	LB136746
	Chromium	2.12	+/-5	U	10.0	P	08/07/2025	13:25	LB136746
	Lead	2.30	+/-6	U	12.0	P	08/07/2025	13:25	LB136746
	Selenium	9.64	+/-10	U	20.0	P	08/07/2025	13:25	LB136746
	Silver	1.62	+/-5	U	10.0	P	08/07/2025	13:25	LB136746

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Sciacca General Contractors, LLC

SDG No.: Q2758

Contract: SCIA01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	08/07/2025	14:02	LB136746
	Barium	14.6	+/-50	U	100	P	08/07/2025	14:02	LB136746
	Cadmium	0.81	+/-3	J	6.00	P	08/07/2025	14:02	LB136746
	Chromium	2.12	+/-5	U	10.0	P	08/07/2025	14:02	LB136746
	Lead	2.30	+/-6	U	12.0	P	08/07/2025	14:02	LB136746
	Selenium	9.64	+/-10	U	20.0	P	08/07/2025	14:02	LB136746
	Silver	1.62	+/-5	U	10.0	P	08/07/2025	14:02	LB136746
CCB02	Arsenic	5.12	+/-10	U	20.0	P	08/07/2025	14:52	LB136746
	Barium	14.6	+/-50	U	100	P	08/07/2025	14:52	LB136746
	Cadmium	1.03	+/-3	J	6.00	P	08/07/2025	14:52	LB136746
	Chromium	2.12	+/-5	U	10.0	P	08/07/2025	14:52	LB136746
	Lead	2.64	+/-6	J	12.0	P	08/07/2025	14:52	LB136746
	Selenium	9.64	+/-10	U	20.0	P	08/07/2025	14:52	LB136746
	Silver	1.62	+/-5	U	10.0	P	08/07/2025	14:52	LB136746
CCB03	Arsenic	5.12	+/-10	U	20.0	P	08/07/2025	15:49	LB136746
	Barium	14.6	+/-50	U	100	P	08/07/2025	15:49	LB136746
	Cadmium	0.50	+/-3	U	6.00	P	08/07/2025	15:49	LB136746
	Chromium	2.12	+/-5	U	10.0	P	08/07/2025	15:49	LB136746
	Lead	2.30	+/-6	U	12.0	P	08/07/2025	15:49	LB136746
	Selenium	9.64	+/-10	U	20.0	P	08/07/2025	15:49	LB136746
	Silver	1.62	+/-5	U	10.0	P	08/07/2025	15:49	LB136746
CCB04	Arsenic	5.12	+/-10	U	20.0	P	08/07/2025	17:00	LB136746
	Barium	14.6	+/-50	U	100	P	08/07/2025	17:00	LB136746
	Cadmium	0.50	+/-3	U	6.00	P	08/07/2025	17:00	LB136746
	Chromium	2.12	+/-5	U	10.0	P	08/07/2025	17:00	LB136746
	Lead	2.30	+/-6	U	12.0	P	08/07/2025	17:00	LB136746
	Selenium	9.64	+/-10	U	20.0	P	08/07/2025	17:00	LB136746
	Silver	1.62	+/-5	U	10.0	P	08/07/2025	17:00	LB136746
CCB05	Arsenic	5.12	+/-10	U	20.0	P	08/07/2025	17:42	LB136746
	Barium	14.6	+/-50	U	100	P	08/07/2025	17:42	LB136746
	Cadmium	0.50	+/-3	U	6.00	P	08/07/2025	17:42	LB136746
	Chromium	2.12	+/-5	U	10.0	P	08/07/2025	17:42	LB136746
	Lead	2.30	+/-6	U	12.0	P	08/07/2025	17:42	LB136746
	Selenium	9.64	+/-10	U	20.0	P	08/07/2025	17:42	LB136746
	Silver	1.62	+/-5	U	10.0	P	08/07/2025	17:42	LB136746