

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Sciacca General Contractors, LLC

SDG No.: Q2758

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB169122TB		WATER		Batch Number:	PB169163		Prep Date:	08/07/2025	
	Mercury	0.76	<2	U	2.00	CV	08/07/2025	15:31	LB136741
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB169163BL		WATER		Batch Number:	PB169163		Prep Date:	08/07/2025	
	Mercury	0.076	<0.2	U	0.20	CV	08/07/2025	14:50	LB136741

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Sciacca General Contractors, LLC

SDG No.: Q2758

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB169122TB		WATER		Batch Number:	PB169142		Prep Date:	08/06/2025	
	Arsenic	25.6	<50	U	100	P	08/07/2025	16:06	LB136746
	Barium	72.8	<250	U	500	P	08/07/2025	16:06	LB136746
	Cadmium	2.50	<15	U	30.0	P	08/07/2025	16:06	LB136746
	Chromium	10.6	<25	U	50.0	P	08/07/2025	16:06	LB136746
	Lead	11.5	<30	U	60.0	P	08/07/2025	16:06	LB136746
	Selenium	48.2	<50	U	100	P	08/07/2025	16:06	LB136746
	Silver	8.10	<25	U	50.0	P	08/07/2025	16:06	LB136746
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB169142BL		WATER		Batch Number:	PB169142		Prep Date:	08/06/2025	
	Arsenic	25.6	<50	U	100	P	08/07/2025	15:57	LB136746
	Barium	72.8	<250	U	500	P	08/07/2025	15:57	LB136746
	Cadmium	2.50	<15	U	30.0	P	08/07/2025	15:57	LB136746
	Chromium	10.6	<25	U	50.0	P	08/07/2025	15:57	LB136746
	Lead	11.5	<30	U	60.0	P	08/07/2025	15:57	LB136746
	Selenium	48.2	<50	U	100	P	08/07/2025	15:57	LB136746
	Silver	8.10	<25	U	50.0	P	08/07/2025	15:57	LB136746