

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Sciacca General Contractors, LLC</u>	SDG No.: <u>Q2758</u>
Contract: <u>SCIA01</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	5.29			-20	20	08/07/2025	13:34	LB136746
	Barium	5.52	6.0	92	-94	106	08/07/2025	13:34	LB136746
	Cadmium	-0.26	1.0	26	-5	7	08/07/2025	13:34	LB136746
	Chromium	59.1	52.0	114	42	62	08/07/2025	13:34	LB136746
	Lead	-8.67			-12	12	08/07/2025	13:34	LB136746
	Selenium	-8.28			-20	20	08/07/2025	13:34	LB136746
	Silver	2.72			-10	10	08/07/2025	13:34	LB136746
ICSAB01	Arsenic	103	104	99	88.4	120	08/07/2025	13:39	LB136746
	Barium	502	537	94	437	637	08/07/2025	13:39	LB136746
	Cadmium	969	972	100	826	1120	08/07/2025	13:39	LB136746
	Chromium	547	542	101	460	624	08/07/2025	13:39	LB136746
	Lead	37.4	49.0	76	37	61	08/07/2025	13:39	LB136746
	Selenium	46.5	46.0	101	26	66	08/07/2025	13:39	LB136746
	Silver	216	201	108	170	232	08/07/2025	13:39	LB136746