

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: SCIA01

Lab Code: ACE

SDG NO.: Q2758

Continuing Calib Date: 08/05/2025

Initial Calibration Date(s): 07/23/2025

07/23/2025

Continuing Calib Time: 16:15

Initial Calibration Time(s): 11:32

19:47

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	4.75	4.76	4.66	4.86	0.01
Aroclor-1242-2	(2)	4.77	4.77	4.67	4.87	0.00
Aroclor-1242-3	(3)	4.83	4.83	4.73	4.93	0.00
Aroclor-1242-4	(4)	4.95	4.95	4.85	5.05	0.00
Aroclor-1242-5	(5)	5.60	5.60	5.50	5.70	0.00
Tetrachloro-m-xylene		3.67	3.67	3.57	3.77	0.00
Decachlorobiphenyl		8.69	8.69	8.59	8.79	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: SCIA01

Lab Code: ACE

SDG NO.: Q2758

Continuing Calib Date: 08/05/2025

Initial Calibration Date(s): 07/23/2025

07/23/2025

Continuing Calib Time: 16:15

Initial Calibration Time(s): 11:32

19:47

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.73	4.74	4.64	4.84	0.01
Aroclor-1242-2 (2)	4.75	4.76	4.66	4.86	0.01
Aroclor-1242-3 (3)	4.93	4.93	4.83	5.03	0.00
Aroclor-1242-4 (4)	5.01	5.02	4.92	5.12	0.01
Aroclor-1242-5 (5)	5.53	5.53	5.43	5.63	0.00
Tetrachloro-m-xylene	3.66	3.66	3.56	3.76	0.00
Decachlorobiphenyl	8.63	8.64	8.54	8.74	0.01