

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: SCIA01

Lab Code: ACE

SDG NO.: Q2758

Continuing Calib Date: 08/05/2025

Initial Calibration Date(s): 07/23/2025 07/23/2025

Continuing Calib Time: 16:52

Initial Calibration Time(s): 11:32 19:47

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.56	5.56	5.46	5.66	0.00
Aroclor-1254-2 (2)	5.71	5.71	5.61	5.81	0.00
Aroclor-1254-3 (3)	6.11	6.11	6.01	6.21	0.00
Aroclor-1254-4 (4)	6.34	6.34	6.24	6.44	0.00
Aroclor-1254-5 (5)	6.76	6.76	6.66	6.86	0.00
Tetrachloro-m-xylene	3.67	3.67	3.57	3.77	0.00
Decachlorobiphenyl	8.69	8.70	8.60	8.80	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: SCIA01

Lab Code: ACE

SDG NO.: Q2758

Continuing Calib Date: 08/05/2025

Initial Calibration Date(s): 07/23/2025

07/23/2025

Continuing Calib Time: 16:52

Initial Calibration Time(s): 11:32

19:47

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	5.53	5.54	5.44	5.64	0.01
Aroclor-1254-2	(2)	5.68	5.68	5.58	5.78	0.00
Aroclor-1254-3	(3)	6.08	6.08	5.98	6.18	0.00
Aroclor-1254-4	(4)	6.31	6.31	6.21	6.41	0.00
Aroclor-1254-5	(5)	6.72	6.73	6.63	6.83	0.01
Tetrachloro-m-xylene		3.66	3.66	3.56	3.76	0.00
Decachlorobiphenyl		8.63	8.64	8.54	8.74	0.01