

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** SCIA01
Lab Code: ACE **SDG NO.:** Q2758
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/05/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112758.D **Time Analyzed:** 15:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.752	4.655	4.855	484.040	500.000	-3.2
Aroclor-1016-2	4.771	4.674	4.874	475.820	500.000	-4.8
Aroclor-1016-3	4.828	4.731	4.931	473.490	500.000	-5.3
Aroclor-1016-4	4.947	4.851	5.051	466.260	500.000	-6.7
Aroclor-1016-5	5.204	5.108	5.308	470.060	500.000	-6.0
Aroclor-1260-1	6.240	6.145	6.345	426.790	500.000	-14.6
Aroclor-1260-2	6.430	6.333	6.533	420.540	500.000	-15.9
Aroclor-1260-3	6.796	6.700	6.900	438.260	500.000	-12.3
Aroclor-1260-4	7.054	6.959	7.159	401.240	500.000	-19.8
Aroclor-1260-5	7.297	7.203	7.403	401.430	500.000	-19.7
Decachlorobiphenyl	8.685	8.593	8.793	38.950	50.000	-22.1
Tetrachloro-m-xylene	3.667	3.569	3.769	47.340	50.000	-5.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** SCIA01
Lab Code: ACE **SDG NO.:** Q2758
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/05/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112758.D **Time Analyzed:** 15:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.733	4.638	4.838	442.110	500.000	-11.6
Aroclor-1016-2	4.751	4.656	4.856	443.630	500.000	-11.3
Aroclor-1016-3	4.926	4.831	5.031	438.510	500.000	-12.3
Aroclor-1016-4	4.969	4.873	5.073	442.050	500.000	-11.6
Aroclor-1016-5	5.180	5.085	5.285	442.470	500.000	-11.5
Aroclor-1260-1	6.208	6.113	6.313	427.230	500.000	-14.6
Aroclor-1260-2	6.397	6.302	6.502	428.600	500.000	-14.3
Aroclor-1260-3	6.547	6.453	6.653	438.450	500.000	-12.3
Aroclor-1260-4	7.017	6.923	7.123	432.480	500.000	-13.5
Aroclor-1260-5	7.259	7.164	7.364	444.070	500.000	-11.2
Decachlorobiphenyl	8.633	8.540	8.740	42.760	50.000	-14.5
Tetrachloro-m-xylene	3.659	3.562	3.762	45.360	50.000	-9.3