

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO072325

Review By	yogesh	Review On	7/23/2025 3:39:31 PM
Supervise By	mohammad	Supervise On	7/25/2025 3:53:10 AM
SubDirectory	PO072325	HP Acquire Method	HP Processing Method PO072325

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO112412.D	23 Jul 2025 10:56		YP/AJ	Ok
2	I.BLK	I.BLK	PO112413.D	23 Jul 2025 11:14		YP/AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PO112414.D	23 Jul 2025 11:32		YP/AJ	Ok
4	AR1660ICC750	AR1660ICC750	PO112415.D	23 Jul 2025 11:50		YP/AJ	Ok
5	AR1660ICC500	AR1660ICC500	PO112416.D	23 Jul 2025 12:08		YP/AJ	Ok
6	AR1660ICC250	AR1660ICC250	PO112417.D	23 Jul 2025 12:27		YP/AJ	Ok,M
7	AR1660ICC050	AR1660ICC050	PO112418.D	23 Jul 2025 12:45		YP/AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PO112419.D	23 Jul 2025 13:03		YP/AJ	Ok
9	AR1232ICC500	AR1232ICC500	PO112420.D	23 Jul 2025 13:22		YP/AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PO112421.D	23 Jul 2025 13:40		YP/AJ	Ok
11	AR1242ICC750	AR1242ICC750	PO112422.D	23 Jul 2025 13:59		YP/AJ	Ok
12	AR1242ICC500	AR1242ICC500	PO112423.D	23 Jul 2025 14:17		YP/AJ	Ok
13	AR1242ICC250	AR1242ICC250	PO112424.D	23 Jul 2025 14:36		YP/AJ	Ok,M
14	AR1242ICC050	AR1242ICC050	PO112425.D	23 Jul 2025 14:54		YP/AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PO112426.D	23 Jul 2025 15:13		YP/AJ	Ok
16	AR1248ICC750	AR1248ICC750	PO112427.D	23 Jul 2025 15:31		YP/AJ	Ok
17	AR1248ICC500	AR1248ICC500	PO112428.D	23 Jul 2025 15:48		YP/AJ	Ok
18	AR1248ICC250	AR1248ICC250	PO112429.D	23 Jul 2025 16:07		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO072325

Review By	yogesh	Review On	7/23/2025 3:39:31 PM
Supervise By	mohammad	Supervise On	7/25/2025 3:53:10 AM
SubDirectory	PO072325	HP Acquire Method	HP Processing Method PO072325
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		

19	AR1248ICC050	AR1248ICC050	PO112430.D	23 Jul 2025 16:25		YP/AJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PO112431.D	23 Jul 2025 16:44		YP/AJ	Ok
21	AR1254ICC750	AR1254ICC750	PO112432.D	23 Jul 2025 17:02		YP/AJ	Ok
22	AR1254ICC500	AR1254ICC500	PO112433.D	23 Jul 2025 17:21		YP/AJ	Ok
23	AR1254ICC250	AR1254ICC250	PO112434.D	23 Jul 2025 17:39		YP/AJ	Ok
24	AR1254ICC050	AR1254ICC050	PO112435.D	23 Jul 2025 17:57		YP/AJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PO112436.D	23 Jul 2025 18:16		YP/AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PO112437.D	23 Jul 2025 18:34		YP/AJ	Ok
27	AR1268ICC750	AR1268ICC750	PO112438.D	23 Jul 2025 18:53		YP/AJ	Ok
28	AR1268ICC500	AR1268ICC500	PO112439.D	23 Jul 2025 19:11		YP/AJ	Ok
29	AR1268ICC250	AR1268ICC250	PO112440.D	23 Jul 2025 19:28		YP/AJ	Ok
30	AR1268ICC050	AR1268ICC050	PO112441.D	23 Jul 2025 19:47		YP/AJ	Ok
31	PO072325ICV500	ICVPO072325	PO112442.D	23 Jul 2025 20:05		YP/AJ	Ok
32	AR1242ICV500	ICVPO072325AR1242	PO112443.D	23 Jul 2025 20:42		YP/AJ	Ok
33	AR1248ICV500	ICVPO072325AR1248	PO112444.D	23 Jul 2025 21:18		YP/AJ	Ok
34	AR1254ICV500	ICVPO072325AR1254	PO112445.D	23 Jul 2025 21:54		YP/AJ	Ok
35	AR1268ICV500	ICVPO072325AR1268	PO112446.D	23 Jul 2025 22:31		YP/AJ	Ok
36	DDT ANALOG	DDT ANALOG	PO112447.D	23 Jul 2025 23:07		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO080525

Review By	yogesh	Review On	8/5/2025 10:39:27 AM
Supervise By		Supervise On	
SubDirectory	PO080525	HP Acquire Method	HP Processing Method PO072325

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO112739.D	05 Aug 2025 08:10		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO112740.D	05 Aug 2025 08:29		YP/AJ	Ok,NS
3	I.BLK	I.BLK	PO112741.D	05 Aug 2025 08:50		YP/AJ	Ok
4	DDT ANALOG	DDT ANALOG	PO112742.D	05 Aug 2025 09:10		YP/AJ	Ok
5	Q2747-01	OU4-TS-GRILLO-OG2	PO112743.D	05 Aug 2025 09:28		YP/AJ	Ok,NS
6	AR1660CCC500	AR1660CCC500	PO112744.D	05 Aug 2025 10:05		YP/AJ	Ok,NS
7	AR1242CCC500	AR1242CCC500	PO112745.D	05 Aug 2025 10:37		YP/AJ	Ok
8	AR1248CCC500	AR1248CCC500	PO112746.D	05 Aug 2025 10:55		YP/AJ	Ok
9	AR1254CCC500	AR1254CCC500	PO112747.D	05 Aug 2025 11:13		YP/AJ	Ok,NS
10	I.BLK	I.BLK	PO112748.D	05 Aug 2025 11:32		YP/AJ	Ok,NS
11	PB169117BL	PB169117BL	PO112749.D	05 Aug 2025 12:00		YP/AJ	Ok
12	PB169117BS	PB169117BS	PO112750.D	05 Aug 2025 12:18		YP/AJ	Ok,NS
13	Q2758-01	WASTE	PO112751.D	05 Aug 2025 12:36		YP/AJ	Ok,NS
14	Q2759-01	SOLID-DRUMS	PO112752.D	05 Aug 2025 12:55		YP/AJ	Ok
15	Q2760-01	NB-07-08042025	PO112753.D	05 Aug 2025 13:12		YP/AJ	Ok,NS
16	Q2760-01MS	NB-07-08042025MS	PO112754.D	05 Aug 2025 13:31		YP/AJ	Ok,NS
17	Q2760-01MSD	NB-07-08042025MSD	PO112755.D	05 Aug 2025 13:49		YP/AJ	Ok,NS
18	Q2763-01	TP-2	PO112756.D	05 Aug 2025 14:07		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO080525

Review By	yogesh	Review On	8/5/2025 10:39:27 AM
Supervise By		Supervise On	
SubDirectory	PO080525	HP Acquire Method	HP Processing Method PO072325
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q2752-03	PIPE-LINE-LIQUIDS	PO112757.D	05 Aug 2025 14:26	TCMX high in both column	YP/AJ	ReRun
20	AR1660CCC500	AR1660CCC500	PO112758.D	05 Aug 2025 15:38	DCB Low in 1st column	YP/AJ	Ok,NS
21	AR1242CCC500	AR1242CCC500	PO112759.D	05 Aug 2025 16:15		YP/AJ	Ok,NS
22	AR1248CCC500	AR1248CCC500	PO112760.D	05 Aug 2025 16:34		YP/AJ	Ok,NS
23	AR1254CCC500	AR1254CCC500	PO112761.D	05 Aug 2025 16:52		YP/AJ	Ok,NS
24	I.BLK	I.BLK	PO112762.D	05 Aug 2025 17:10		YP/AJ	Ok,NS

M : Manual Integration