



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q2758

Order ID: Q2758

Client: Sciacca General Contractors, LLC

Project ID: 92 Clifford Street, Newark

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : Q2758-01	WASTE WASTE	SOIL	Aroclor-1260	32.5	3.50	18.3	ug/kg	
Total Concentration:				32.500				



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	Sciacca General Contractors, LLC		Date Collected:	08/01/25	
Project:	92 Clifford Street, Newark		Date Received:	08/01/25	
Client Sample ID:	WASTE		SDG No.:	Q2758	
Lab Sample ID:	Q2758-01		Matrix:	SOIL	
Analytical Method:	8082A		% Solid:	92.9	Decanted:
Sample Wt/Vol:	30.06	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO112751.D	1	08/05/25 08:20	08/05/25 12:36	PB169117

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.20	U	4.20	18.3	ug/kg
11104-28-2	Aroclor-1221	4.30	U	4.30	18.3	ug/kg
11141-16-5	Aroclor-1232	4.00	U	4.00	18.3	ug/kg
53469-21-9	Aroclor-1242	4.30	U	4.30	18.3	ug/kg
12672-29-6	Aroclor-1248	6.40	U	6.40	18.3	ug/kg
11097-69-1	Aroclor-1254	3.40	U	3.40	18.3	ug/kg
37324-23-5	Aroclor-1262	5.40	U	5.40	18.3	ug/kg
11100-14-4	Aroclor-1268	3.90	U	3.90	18.3	ug/kg
11096-82-5	Aroclor-1260	32.5		3.50	18.3	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.0		32 - 144	85%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.0		32 - 175	90%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

LAB CHRONICLE

OrderID:	Q2758	OrderDate:	8/1/2025 4:14:00 PM
Client:	Sciacca General Contractors, LLC	Project:	92 Clifford Street, Newark
Contact:	Rosanne Scirica	Location:	D21,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2758-01	WASTE	SOIL	PCB	8082A	08/01/25	08/05/25	08/05/25	08/01/25