

Surrogate Summary

SDG No.: Q2758

Client: Sciacca General Contractors, LLC

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO112413.D	PIBLK-PO112413.D	Tetrachloro-m-xyl	1	20	19.1	95		60	140
		Decachlorobiphen	1	20	20.0	100		60	140
		Tetrachloro-m-xyl	2	20	19.5	98		60	140
		Decachlorobiphen	2	20	20.4	102		60	140
I.BLK-PO112748.D	PIBLK-PO112748.D	Tetrachloro-m-xyl	1	20	18.1	91		60	140
		Decachlorobiphen	1	20	17.8	89		60	140
		Tetrachloro-m-xyl	2	20	17.4	87		60	140
		Decachlorobiphen	2	20	18.3	91		60	140
PB169117BL	PB169117BL	Tetrachloro-m-xyl	1	20	21.1	105		32	144
		Decachlorobiphen	1	20	21.0	105		32	175
		Tetrachloro-m-xyl	2	20	19.9	99		32	144
		Decachlorobiphen	2	20	21.2	106		32	175
PB169117BS	PB169117BS	Tetrachloro-m-xyl	1	20	20.1	100		32	144
		Decachlorobiphen	1	20	20.5	103		32	175
		Tetrachloro-m-xyl	2	20	19.0	95		32	144
		Decachlorobiphen	2	20	20.7	104		32	175
Q2758-01	WASTE	Tetrachloro-m-xyl	1	20	17.0	85		32	144
		Decachlorobiphen	1	20	16.0	80		32	175
		Tetrachloro-m-xyl	2	20	16.8	84		32	144
		Decachlorobiphen	2	20	18.0	90		32	175
Q2760-01MS	NB-07-08042025MS	Tetrachloro-m-xyl	1	20	20.3	101		30	150
		Decachlorobiphen	1	20	15.1	76		30	150
		Tetrachloro-m-xyl	2	20	19.4	97		30	150
		Decachlorobiphen	2	20	18.1	90		30	150
Q2760-01MSD	NB-07-08042025MSD	Tetrachloro-m-xyl	1	20	19.8	99		30	150
		Decachlorobiphen	1	20	14.5	73		30	150
		Tetrachloro-m-xyl	2	20	19.0	95		30	150
		Decachlorobiphen	2	20	17.3	86		30	150
I.BLK-PO112762.D	PIBLK-PO112762.D	Tetrachloro-m-xyl	1	20	17.8	89		60	140
		Decachlorobiphen	1	20	16.4	82		60	140
		Tetrachloro-m-xyl	2	20	17.4	87		60	140
		Decachlorobiphen	2	20	17.5	87		60	140