

**TPH GC CONTINUING CALIBRATION SUMMARY**

**50 PPM TRPH STD**

Lab Name: Alliane Contract: SCIA01  
ProjectID: 92 Clifford Street, Newark  
Lab Code: ACE SDG No.: Q2758  
DataFile: FG016363.D Analyst Name: YP\AJ Analyst Date: 08-06-2025

| Conc. (PPM) | Area Count | RF     | Average RF | %D    |
|-------------|------------|--------|------------|-------|
| 850         | 87118415   | 102492 | 99114      | 3.408 |