

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BF143320.D
Instrument ID: BNA_F

Contract: SCIA01
SDG NO.: Q2758
DFTPP Injection Date: 08/06/2025
DFTPP Injection Time: 09:39

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.6 (1.9) 1
69	Mass 69 relative abundance	29
70	Less than 2.0% of mass 69	0.1 (0.4) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.3
365	Greater than 1% of mass 198	2.8
441	Present, but less than mass 443	15.1
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.5 (19.5) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF143321.D	08/06/2025	10:38
PB169119BL	PB169119BL	BF143324.D	08/06/2025	12:05
PB169119BS	PB169119BS	BF143325.D	08/06/2025	12:34
TP-2MS	Q2763-01MS	BF143327.D	08/06/2025	13:49
TP-2MSD	Q2763-01MSD	BF143328.D	08/06/2025	14:18