

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2758

Analytical Method: SW8270E

Client: Sciaccia General Contractors, LLC

DataFile: BF143327.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: Q2763-01MS		Client Sample ID: TP-2MS									
Naphthalene	1200	0	900	ug/Kg	75				51	121	
Acenaphthylene	1200	0	930	ug/Kg	78				57	125	
Acenaphthene	1200	65.7	1100	ug/Kg	86				70	121	
Fluorene	1200	68.5	1100	ug/Kg	86				53	118	
Phenanthrene	1200	370	1100	ug/Kg	61				52	128	
Anthracene	1200	100	940	ug/Kg	70				62	124	
Fluoranthene	1200	970	1400	ug/Kg	36	*			44	125	
Pyrene	1200	410	900	ug/Kg	41				37	122	
Benzo(a)anthracene	1200	400	1100	ug/Kg	58				53	119	
Chrysene	1200	340	1100	ug/Kg	63				57	121	
Benzo(b)fluoranthene	1200	490	1400	ug/Kg	76				52	117	
Benzo(k)fluoranthene	1200	230	1400	ug/Kg	98				57	134	
Benzo(a)pyrene	1200	360	1200	ug/Kg	70				70	142	
Indeno(1,2,3-cd)pyrene	1200	140	810	ug/Kg	56				40	129	
Dibenz(a,h)anthracene	1200	0	730	ug/Kg	61				43	123	
Benzo(g,h,i)perylene	1200	190	720	ug/Kg	44				24	125	