

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2758

Analytical Method: SW8270E

Client: Sciacca General Contractors, LLC

DataFile: BF143328.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: Q2763-01MSD		Client Sample ID: TP-2MSD									
Naphthalene	1200	0	1000	ug/Kg	83		10		51	121	20
Acenaphthylene	1200	0	1100	ug/Kg	92		16		57	125	20
Acenaphthene	1200	65.7	1100	ug/Kg	86		0		70	121	20
Fluorene	1200	68.5	1100	ug/Kg	86		0		53	118	20
Phenanthrene	1200	370	1300	ug/Kg	78		24	*	52	128	20
Anthracene	1200	100	1100	ug/Kg	83		17		62	124	20
Fluoranthene	1200	970	1700	ug/Kg	61		52	*	44	125	20
Pyrene	1200	410	1000	ug/Kg	49		18		37	122	20
Benzo(a)anthracene	1200	400	1300	ug/Kg	75		26	*	53	119	20
Chrysene	1200	340	1200	ug/Kg	72		13		57	121	20
Benzo(b)fluoranthene	1200	490	1700	ug/Kg	101		28	*	52	117	20
Benzo(k)fluoranthene	1200	230	1300	ug/Kg	89		10		57	134	20
Benzo(a)pyrene	1200	360	1300	ug/Kg	78		11		70	142	20
Indeno(1,2,3-cd)pyrene	1200	140	920	ug/Kg	65		15		40	129	20
Dibenz(a,h)anthracene	1200	0	810	ug/Kg	68		11		43	123	20
Benzo(g,h,i)perylene	1200	190	950	ug/Kg	63		36	*	24	125	20