

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: <u>Alliance</u>	Contract: <u>SCIA01</u>
Lab Code: <u>ACE</u>	SDG No.: <u>Q2758</u>
Instrument ID: <u>MSVOA_W</u>	Calibration Date(s): <u>07/22/2025</u> <u>07/22/2025</u>
Heated Purge: (Y/N) <u>Y</u>	Calibration Time(s): <u>09:05</u> <u>12:00</u>
GC Column: <u>RXI-624</u> ID: <u>0.25</u> (mm)	

LAB FILE ID:		RRF005 = VW031891.D		RRF010 = VW031892.D		RRF020 = VW031893.D		
		RRF050 = VW031894.D		RRF100 = VW031895.D		RRF150 = VW031896.D		
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
Dichlorodifluoromethane	0.427	0.364	0.355	0.358	0.369	0.357	0.372	7.4
Chloromethane	0.533	0.475	0.462	0.434	0.459	0.498	0.477	7.2
Vinyl Chloride	0.635	0.615	0.612	0.615	0.617	0.624	0.620	1.3
Bromomethane	0.500	0.454	0.444	0.434	0.448	0.457	0.456	5
Chloroethane	0.448	0.395	0.394	0.392	0.409	0.414	0.409	5.2
Trichlorofluoromethane	0.670	0.552	0.574	0.679	0.625	0.666	0.628	8.6
1,1,2-Trichlorotrifluoroethane	0.674	0.603	0.582	0.589	0.584	0.590	0.604	5.8
1,1-Dichloroethene	0.714	0.656	0.650	0.645	0.667	0.669	0.667	3.7
Acetone	0.206	0.189	0.186	0.166	0.168	0.163	0.179	9.3
Carbon Disulfide	1.775	1.643	1.708	1.693	1.757	1.798	1.729	3.4
Methyl tert-butyl Ether	1.271	1.110	1.204	1.132	1.153	1.042	1.152	6.9
Methyl Acetate	0.528	0.427	0.529	0.475	0.527	0.481	0.494	8.4
Methylene Chloride	1.615	0.931	0.910	0.756	0.730	0.722	0.944	36.1
trans-1,2-Dichloroethene	0.767	0.679	0.697	0.691	0.699	0.712	0.707	4.4
1,1-Dichloroethane	1.372	1.262	1.274	1.263	1.291	1.296	1.293	3.2
Cyclohexane	1.399	1.176	1.078	1.099	1.100	1.107	1.160	10.5
2-Butanone	0.260	0.231	0.266	0.236	0.254	0.227	0.246	6.7
Carbon Tetrachloride	0.489	0.498	0.502	0.493	0.517	0.506	0.501	2
cis-1,2-Dichloroethene	0.878	0.789	0.821	0.824	0.846	0.831	0.832	3.6
Bromochloromethane	0.585	0.520	0.546	0.526	0.542	0.531	0.541	4.3
Chloroform	1.436	1.336	1.362	1.314	1.348	1.338	1.356	3.1
1,1,1-Trichloroethane	1.114	1.017	1.008	1.010	1.009	1.032	1.032	4
Methylcyclohexane	0.655	0.643	0.637	0.657	0.690	0.659	0.657	2.8
Benzene	1.558	1.540	1.565	1.520	1.569	1.517	1.545	1.5
1,2-Dichloroethane	0.508	0.484	0.516	0.480	0.507	0.471	0.494	3.7
Trichloroethene	0.370	0.372	0.356	0.362	0.375	0.368	0.367	1.8
1,2-Dichloropropane	0.381	0.367	0.376	0.360	0.377	0.370	0.372	2.1
Bromodichloromethane	0.537	0.546	0.561	0.535	0.582	0.557	0.553	3.2
4-Methyl-2-Pentanone	0.293	0.270	0.317	0.295	0.320	0.273	0.295	7.1
Toluene	0.992	0.998	1.003	0.979	0.998	0.986	0.993	0.9

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>Alliance</u>	Contract:	<u>SCIA01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2758</u>
Instrument ID:	<u>MSVOA_W</u>	Calibration Date(s):	<u>07/22/2025</u> <u>07/22/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Calibration Time(s):	<u>09:05</u> <u>12:00</u>
GC Column:	<u>RXI-624</u> ID: <u>0.25</u> (mm)		

LAB FILE ID:		RRF005 = VW031891.D		RRF010 = VW031892.D		RRF020 = VW031893.D		
		RRF050 = VW031894.D		RRF100 = VW031895.D		RRF150 = VW031896.D		
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
t-1,3-Dichloropropene	0.502	0.496	0.535	0.522	0.575	0.552	0.530	5.7
cis-1,3-Dichloropropene	0.591	0.578	0.601	0.602	0.646	0.617	0.606	3.9
1,1,2-Trichloroethane	0.316	0.297	0.314	0.300	0.317	0.295	0.306	3.3
2-Hexanone	0.205	0.188	0.220	0.206	0.222	0.189	0.205	7.1
Dibromochloromethane	0.349	0.339	0.358	0.351	0.386	0.359	0.357	4.5
1,2-Dibromoethane	0.302	0.291	0.318	0.297	0.318	0.290	0.303	4.2
Tetrachloroethene	0.345	0.334	0.333	0.324	0.319	0.323	0.330	2.9
Chlorobenzene	1.216	1.227	1.257	1.244	1.193	1.205	1.224	2
Ethyl Benzene	2.167	2.107	2.165	2.226	2.099	2.137	2.150	2.2
m/p-Xylenes	0.808	0.801	0.829	0.823	0.794	0.802	0.809	1.7
o-Xylene	0.734	0.742	0.774	0.796	0.769	0.780	0.766	3.1
Styrene	1.277	1.260	1.344	1.361	1.276	1.308	1.304	3.1
Bromoform	0.202	0.182	0.208	0.219	0.223	0.211	0.207	7.1
Isopropylbenzene	4.192	4.207	4.224	4.405	4.614	4.387	4.338	3.8
1,1,2,2-Tetrachloroethane	0.947	0.851	0.964	0.914	0.982	0.868	0.921	5.7
1,3-Dichlorobenzene	2.060	1.901	2.017	1.911	1.977	1.858	1.954	3.9
1,4-Dichlorobenzene	2.044	1.907	1.985	1.824	1.942	1.783	1.914	5.1
1,2-Dichlorobenzene	1.806	1.694	1.765	1.677	1.802	1.646	1.732	3.9
1,2-Dibromo-3-Chloropropane	0.183	0.145	0.169	0.160	0.180	0.160	0.166	8.5
1,2,4-Trichlorobenzene	1.113	0.952	1.036	0.984	1.171	1.010	1.044	7.9
1,2,3-Trichlorobenzene	0.976	0.898	0.907	0.935	1.042	0.920	0.946	5.7
1,2-Dichloroethane-d4	0.899	0.744	0.782	0.709	0.717	0.686	0.756	10.2
Dibromofluoromethane	0.358	0.353	0.346	0.326	0.330	0.324	0.340	4.3
Toluene-d8	1.393	1.309	1.327	1.224	1.259	1.204	1.286	5.5
4-Bromofluorobenzene	0.542	0.515	0.516	0.460	0.476	0.452	0.494	7.2

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.