

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: SCIA01
 Lab Code: ACE SDG No.: Q2758
 Instrument ID: MSVOA_W Calibration Date/Time: 08/04/2025 09:24
 Lab File ID: VW031987.D Init. Calib. Date(s): 07/22/2025 07/22/2025
 Heated Purge: (Y/N) Y Init. Calib. Time(s): 09:05 12:00
 GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Methylene Chloride	50.000	45.27	9.470001	20