

Surrogate Summary

SDG No.: Q2758

Client: Sciacca General Contractors, LLC

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2758-01	WASTE	1,2-Dichloroethane-d4	50	51.9	104		63	155
		Dibromofluoromethane	50	46.8	94		70	134
		Toluene-d8	50	45.5	91		74	123
		4-Bromofluorobenzene	50	45.0	90		17	146
VW0804SBL01	VW0804SBL01	1,2-Dichloroethane-d4	50	50.6	101		63	155
		Dibromofluoromethane	50	48.1	96		70	134
		Toluene-d8	50	45.3	91		74	123
		4-Bromofluorobenzene	50	41.6	83		17	146
VW0804SBS01	VW0804SBS01	1,2-Dichloroethane-d4	50	52.8	106		63	155
		Dibromofluoromethane	50	55.9	112		70	134
		Toluene-d8	50	55.6	111		74	123
		4-Bromofluorobenzene	50	55.9	112		17	146
VW0804SBSD0	VW0804SBSD01	1,2-Dichloroethane-d4	50	51.4	103		63	155
		Dibromofluoromethane	50	50.9	102		70	134
		Toluene-d8	50	49.5	99		74	123
		4-Bromofluorobenzene	50	48.6	97		17	146