

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2758</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Sciacca General Contractors, LLC</u>	Datafile :	<u>VW031990.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VW0804SBSD01	Dichlorodifluoromethane	20	16.5	ug/Kg	83	4		64	136	20
	Chloromethane	20	18.9	ug/Kg	95	3		52	151	20
	Vinyl chloride	20	17.1	ug/Kg	86	3		56	148	20
	Bromomethane	20	19.1	ug/Kg	96	2		58	141	20
	Chloroethane	20	19.6	ug/Kg	98	2		69	130	20
	Trichlorofluoromethane	20	17.3	ug/Kg	86	12		69	134	20
	1,1,2-Trichlorotrifluoroethane	20	18.9	ug/Kg	95	4		81	123	20
	1,1-Dichloroethene	20	18.3	ug/Kg	92	3		79	121	20
	Acetone	100	120	ug/Kg	120	0		40	171	20
	Carbon disulfide	20	18.0	ug/Kg	90	3		59	130	20
	Methyl tert-butyl Ether	20	20.2	ug/Kg	101	1		77	129	20
	Methyl Acetate	20	20.7	ug/Kg	104	5		69	149	20
	Methylene Chloride	20	16.1	ug/Kg	81	2		72	131	20
	trans-1,2-Dichloroethene	20	19.1	ug/Kg	96	0		80	123	20
	1,1-Dichloroethane	20	19.6	ug/Kg	98	0		82	123	20
	Cyclohexane	20	18.3	ug/Kg	92	2		76	122	20
	2-Butanone	100	110	ug/Kg	110	10		69	131	20
	Carbon Tetrachloride	20	18.1	ug/Kg	91	10		76	129	20
	cis-1,2-Dichloroethene	20	19.8	ug/Kg	99	1		82	123	20
	Bromochloromethane	20	19.1	ug/Kg	96	4		80	127	20
	Chloroform	20	20.1	ug/Kg	101	1		82	125	20
	1,1,1-Trichloroethane	20	19.4	ug/Kg	97	4		80	126	20
	Methylcyclohexane	20	17.4	ug/Kg	87	7		77	123	20
	Benzene	20	18.8	ug/Kg	94	8		84	121	20
	1,2-Dichloroethane	20	19.2	ug/Kg	96	4		81	126	20
	Trichloroethene	20	19.3	ug/Kg	97	5		83	122	20
	1,2-Dichloropropane	20	19.3	ug/Kg	97	4		83	122	20
	Bromodichloromethane	20	19.3	ug/Kg	97	6		82	123	20
	4-Methyl-2-Pentanone	100	98.1	ug/Kg	98	2		70	135	20
	Toluene	20	18.7	ug/Kg	94	7		83	122	20
	t-1,3-Dichloropropene	20	18.8	ug/Kg	94	3		78	124	20
	cis-1,3-Dichloropropene	20	19.0	ug/Kg	95	6		81	122	20
	1,1,2-Trichloroethane	20	20.3	ug/Kg	102	0		82	125	20
	2-Hexanone	100	100	ug/Kg	100	1		66	138	20
	Dibromochloromethane	20	18.5	ug/Kg	93	10		79	125	20
	1,2-Dibromoethane	20	19.4	ug/Kg	97	2		80	125	20
	Tetrachloroethene	20	18.7	ug/Kg	94	9		83	125	20
	Chlorobenzene	20	18.9	ug/Kg	95	3		84	122	20
	Ethyl Benzene	20	18.3	ug/Kg	92	5		82	124	20
	m/p-Xylenes	40	38.2	ug/Kg	96	3		83	124	20
	o-Xylene	20	18.6	ug/Kg	93	6		83	123	20
	Styrene	20	19.3	ug/Kg	97	4		82	124	20
	Bromoform	20	19.9	ug/Kg	100	3		75	127	20
	Isopropylbenzene	20	17.9	ug/Kg	90	2		82	124	20
	1,1,2,2-Tetrachloroethane	20	19.2	ug/Kg	96	3		77	127	20
	1,3-Dichlorobenzene	20	18.2	ug/Kg	91	6		83	122	20
	1,4-Dichlorobenzene	20	19.1	ug/Kg	96	1		84	121	20
	1,2-Dichlorobenzene	20	19.5	ug/Kg	98	1		83	124	20



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2758 Analytical Method: SW8260D
Client: Sciacca General Contractors, LLC Datafile : VW031990.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VW0804SBSD01	1,2-Dibromo-3-Chloropropane	20	18.9	ug/Kg	95	10		66	134	20
	1,2,4-Trichlorobenzene	20	18.8	ug/Kg	94	5		78	127	20
	1,2,3-Trichlorobenzene	20	19.5	ug/Kg	98	3		70	137	20