

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: SCIA01
 Lab Code: ACE SDG NO.: Q2758
 Lab File ID: VW031987.D Date Analyzed: 08/04/2025
 Instrument ID: MSVOA_W Time Analyzed: 09:24
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	244933	7.96	432425	8.85	389959	11.63
UPPER LIMIT	489866	8.459	864850	9.349	779918	12.129
LOWER LIMIT	122467	7.459	216213	8.349	194980	11.129
EPA SAMPLE NO.						
WASTE	181912	7.97	397326	8.86	370242	11.63
VW0804SBL01	168222	7.95	365101	8.85	334575	11.63
VW0804SBS01	212167	7.96	385271	8.85	349774	11.63
VW0804SBSD01	219677	7.96	423534	8.86	380864	11.63

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.