

SOP ID : MSM4500-NH3 B,G-Ammonia-18

SDG No : N/A

Start Digest Date: 08/05/2025 Time : 14:20 Temp : 150 °C

Matrix : WATER

End Digest Date: 08/05/2025 Time : 15:20 Temp : 157 °C

Pipette ID : WC

Balance ID : N/A

Hood ID : HOOD#2

Digestion tube ID : M5595

Block Thermometer ID : WC CYANIDE

Block ID : WC-DIST-BLOCK-1

Filter paper ID : N/A

Prep Technician Signature: RM

Weigh By : RM

pH Meter ID : N/A

Supervisor Signature: 12

Standardized Name	MLS USED	STD REF. # FROM LOG
LCSW	1.0ML	WP113889
MS/MSD SPIKE SOL.	1.0ML	WP113888
PBW	50.0ML	W3112
RL CHECK	0.1ML	WP113888
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
BORATE BUFFER	2.5ML	WP113886
NAOH 6N	0.5-2.0ML	WP113887
H2SO4 0.04N	5.0ML	WP112828
pH strip-Ammonia	N/A	W3133
KI-starch paper	N/A	W3155
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

ALL GLASSWEAR ARE STEAMED OUT AND THERE WERE NO TRACE OF AMMONIA USING NESLER REAGENT
WP114104,

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
08/05/2025 15:35	RM (WC)	RM (WC)
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	Initial Vol (ml)	Final Vol (ml)	pH	Sulfide	Oxidizing	Nitrate/ Nitrite	Comment	Prep Pos
PB169130BL	PBW130	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
PB169130BS	LCS130	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q2759-04	LIQUID-DRUMS	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q2769-01	001-WILLETS-PT-BLVD(AUG)	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q2769-04	002-35TH-AVE(AUG)	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q2769-04DUP	002-35TH-AVE(AUG)DUP	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q2769-04MS	002-35TH-AVE(AUG)MS	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q2769-04MSD	002-35TH-AVE(AUG)MSD	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A