

Initial and Continuing Calibration Verification

Client: Tully Environmental, Inc

SDG No.: Q2769

Project: Transfer Station-SPDES

RunNo.: LB136726

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	0.98	1	98	90-110	08/06/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.97	1	97	90-110	08/06/2025
Sample ID: CCV2 Ammonia as N	mg/L	1	1	100	90-110	08/06/2025
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	08/06/2025